


LB131045

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.0821e-004 slope

C= -6.3351e-003 = intercept

Rho= 0.9999041

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.00	0.000	0.010	0.010	79	0.000	79					
0.05	0.050					442	0				=
0.20	0.200	0.259	0.059	1276	0.0 %	1276					30
2.50	2.500	2.417	-0.083	11641	0.0 %	11641					13
5.00	5.000	4.960	-0.040	23854	0.0 %	23854					11
7.50	7.500	7.560	0.060	36340	0.0 %	36340					11
10.0	10.000	9.993	-0.007	48023	0.0 %	48023					0

LB131045

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	79	-		0 SFAM01.1	PPB	6/4/2024 9:19	Std
	0.2 S01	1276	-		0.2 SFAM01.1	PPB	6/4/2024 9:24	Std
	2.5 S02	11641	-		2.5 SFAM01.1	PPB	6/4/2024 9:26	Std
	5 S03	23854	-		5 SFAM01.1	PPB	6/4/2024 9:28	Std
	7.5 S04	36340	-		7.5 SFAM01.1	PPB	6/4/2024 9:32	Std
	10 S05	48023	-		10 SFAM01.1	PPB	6/4/2024 9:35	Std
ICV073	ICV073	18408	3.8264 -		SFAM01.1	PPB	6/4/2024 9:37	SMPL
ICB073	ICB073	92	0.0128 -		SFAM01.1	PPB	6/4/2024 9:40	SMPL
CCV076	CCV076	24240	5.0407 -		SFAM01.1	PPB	6/4/2024 9:42	SMPL
CCB076	CCB076	-99	-0.0269 -		SFAM01.1	PPB	6/4/2024 9:44	SMPL
PB161298BL	PBS298	0	-0.0063 -		SFAM01.1	PPB	6/4/2024 9:46	SMPL
P2594-01	MBH6E7	3692	0.7624 -		SFAM01.1	PPB	6/4/2024 9:49	SMPL
P2594-02	MBHBK8	6085	1.2606 -		SFAM01.1	PPB	6/4/2024 9:51	SMPL
P2594-03	MBHBK9	20481	4.2581 -		SFAM01.1	PPB	6/4/2024 9:53	SMPL
P2594-04	MBHBL0	6603	1.3685 -		SFAM01.1	PPB	6/4/2024 9:55	SMPL
P2594-05	MBHBL1	6618	1.3716 -		SFAM01.1	PPB	6/4/2024 9:58	SMPL
P2594-06	MBHBL2	8141	1.6887 -		SFAM01.1	PPB	6/4/2024 10:00	SMPL
P2594-07	MBHBL3	4089	0.845 -		SFAM01.1	PPB	6/4/2024 10:02	SMPL
P2594-08	MBHBL4	5614	1.1626 -		SFAM01.1	PPB	6/4/2024 10:05	SMPL
P2594-09	MBHBL5	7029	1.4572 -		SFAM01.1	PPB	6/4/2024 10:07	SMPL
P2594-10	MBHBL6	3778	0.7803 -		SFAM01.1	PPB	6/4/2024 10:09	SMPL
P2594-11	MBHBL7	146608	30.5192 -		SFAM01.1	PPB	6/4/2024 10:11	SMPL
P2594-12	MBHBL8	2781	0.5727 -		SFAM01.1	PPB	6/4/2024 10:14	SMPL
P2594-13	MBHBL9	5062	1.0476 -		SFAM01.1	PPB	6/4/2024 10:16	SMPL
P2594-14	MBHBM0	7522	1.5598 -		SFAM01.1	PPB	6/4/2024 10:18	SMPL
P2594-15	MBHBM0D	7555	1.5667 -		SFAM01.1	PPB	6/4/2024 10:21	SMPL
P2594-16	MBHBM0S	19195	3.9903 -		SFAM01.1	PPB	6/4/2024 10:23	SMPL
P2594-17	MBHBM1	3551	0.733 -		SFAM01.1	PPB	6/4/2024 10:25	SMPL
P2594-18	MBHBM2	17629	3.6642 -		SFAM01.1	PPB	6/4/2024 10:28	SMPL
P2594-19	MBHBM3	5022	1.0393 -		SFAM01.1	PPB	6/4/2024 10:30	SMPL
CCV77	CCV77	24167	5.0255 -		SFAM01.1	PPB	6/4/2024 10:32	SMPL
CCB77	CCB77	-78	-0.0226 -		SFAM01.1	PPB	6/4/2024 10:34	SMPL
P2594-20	MBHBM4	3257	0.6718 -		SFAM01.1	PPB	6/4/2024 10:37	SMPL
P2594-21	MBHBM5	6962	1.4432 -		SFAM01.1	PPB	6/4/2024 10:39	SMPL
P2594-22	MBHBM6	7096	1.4711 -		SFAM01.1	PPB	6/4/2024 10:41	SMPL
PB161299BL	PBS299	-39	-0.0145 -		SFAM01.1	PPB	6/4/2024 10:43	SMPL
P2587-01	MA4D14	4084	0.844 -		SFAM01.1	PPB	6/4/2024 10:46	SMPL
P2587-02	MA4D15	3013	0.621 -		SFAM01.1	PPB	6/4/2024 10:48	SMPL
P2587-03	MA4D16	82100	17.0879 -		SFAM01.1	PPB	6/4/2024 10:50	SMPL
P2587-04	MA4D17	84746	17.6388 -		SFAM01.1	PPB	6/4/2024 10:53	SMPL
P2587-05	MA4D18	18109	3.7642 -		SFAM01.1	PPB	6/4/2024 10:55	SMPL
P2587-06	MA4D18D	15643	3.2507 -		SFAM01.1	PPB	6/4/2024 10:57	SMPL
P2587-07	MA4D18S	28708	5.971 -		SFAM01.1	PPB	6/4/2024 11:00	SMPL
P2587-08	MA4D19	3636	0.7507 -		SFAM01.1	PPB	6/4/2024 11:02	SMPL
P2587-09	MA4D20	3089	0.6368 -		SFAM01.1	PPB	6/4/2024 11:04	SMPL
P2587-10	MA4D21	5176	1.0714 -		SFAM01.1	PPB	6/4/2024 11:06	SMPL

LB131045

INSTRUMENT ID : CV1

P2587-11	MA4D30	1308	0.266 -	SFAM01.1	PPB	6/4/2024 11:09 SMPL
P2587-12	MA4C28	1138	0.2306 -	SFAM01.1	PPB	6/4/2024 11:11 SMPL
P2587-13	MA4C29	1247	0.2533 -	SFAM01.1	PPB	6/4/2024 11:13 SMPL
P2587-14	MA4C31	1358	0.2764 -	SFAM01.1	PPB	6/4/2024 11:16 SMPL
P2587-15	MA4C37	399	0.0767 -	SFAM01.1	PPB	6/4/2024 11:18 SMPL
P2587-16	MA4C38	351	0.0667 -	SFAM01.1	PPB	6/4/2024 11:20 SMPL
CCV078	CCV078	22693	4.7186 -	SFAM01.1	PPB	6/4/2024 11:22 SMPL
CCB078	CCB078	-183	-0.0444 -	SFAM01.1	PPB	6/4/2024 11:25 SMPL
P2587-17	MA4C44	359	0.0684 -	SFAM01.1	PPB	6/4/2024 11:27 SMPL
P2587-18	MA4C56	1703	0.3483 -	SFAM01.1	PPB	6/4/2024 11:29 SMPL
P2594-11DLX5	MBHBL7	28482	5.924 -	SFAM01.1	PPB	6/4/2024 11:31 SMPL
P2587-03DLX5	MA4D16	15461	3.2128 -	SFAM01.1	PPB	6/4/2024 11:34 SMPL
P2587-04DLX5	MA4D17	15676	3.2576 -	SFAM01.1	PPB	6/4/2024 11:39 SMPL
CCV079	CCV079	23751	4.9389 -	SFAM01.1	PPB	6/4/2024 11:41 SMPL
CCB079	CCB079	-154	-0.0384 -	SFAM01.1	PPB	6/4/2024 11:43 SMPL