

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131056

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	06/04/2024	12:00
2	P2688-01		1	11	100.00	>212.0	>212.0	06/04/2024	12:30
3	P2688-01DUP		1	11	100.00	>212.0	>212.0	06/04/2024	13:45
4	P2691-01		1	8	59.00	138.2	137.9	06/04/2024	14:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6/13/056

WorkList Name : fp6-4

WorkList ID : 180774

Department : Wet-Chemistry

Date : 06-04-2024 10:20:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2688-01	OIL	Solid	Flash Point	Cool 4 deg C	PSEG03	K21	05/31/2024	1010B
P2691-01	DIESEL-1	Solid	Flash Point	Cool 4 deg C	PSEG03	K21	05/31/2024	1010B

Date/Time 06/04/2024 11:50
Raw Sample Received by: RTH CWC
Raw Sample Relinquished by: RTH CWC

Date/Time 06/04/2024 14:35
Raw Sample Received by: RTH CWC
Raw Sample Relinquished by: RTH CWC