

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131062

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	06/05/2024	09:50
2	P2691-02		1	12	100.00	>212.0	>212.0	06/05/2024	10:20
3	P2691-02DUP		1	12	100.00	>212.0	>212.0	06/05/2024	10:50
4	P2691-03		1	14	100.00	>212.0	>212.0	06/05/2024	11:20
5	P2691-04		1	12	100.00	>212.0	>212.0	06/05/2024	11:50
6	P2714-01		1	11	100.00	>212.0	>212.0	06/05/2024	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66131062

WorkList Name : fp-6-3

WorkList ID : 180734

Department : Wet-Chemistry

Date : 06-03-2024 08:25:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2691-02	RINSATE-1	Water	Flash Point	Cool 4 deg C	PSEG03	K21	05/31/2024	1010B
P2691-03	RINSATE-2	Water	Flash Point	Cool 4 deg C	PSEG03	K21	05/31/2024	1010B
P2691-04	50362	Water	Flash Point	Cool 4 deg C	PSEG03	K21	05/31/2024	1010B
P2714-01	N28660	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/04/2024	1010B

Date/Time 06/05/2024 09:45
 Raw Sample Received by: RM CWS
 Raw Sample Relinquished by: DP CWS

Date/Time 06/05/2024 13:00
 Raw Sample Received by: DP CWS
 Raw Sample Relinquished by: RM CWS