

LB131063


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39148



A= 0.0000e+000

B= 2.5594e-004

C= -1.0240e-001

Rho= 0.9996874

Accept=Accepted

Max. Conc.: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% 17
0.0	0.000	-0.046	-0.046	222	0.000	222	0				-16
0.2	0.200	0.169	-0.031	1059	0.0 %	1059					-;
2.5	2.500	2.483	-0.017	10100	0.0 %	10100					4
5.0	5.000	5.201	0.201	20720	0.0 %	20720					0
7.5	7.500	7.476	-0.024	29611	0.0 %	29611					-7
10.0	10.000	9.917	-0.083	39148	0.0 %	39148					

LB131063 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Jun 2024 11:39:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	222	-	0.0000	7470A	PPB	05 Jun 2024 11:54:53	S	Std
0.2 - 1	50.2	1059	-	0.2000	7470A	PPB	05 Jun 2024 11:57:10	S	Std
2.5 - 1	52.5	10100	-	2.5000	7470A	PPB	05 Jun 2024 11:59:27	S	Std
5.0 - 1	55	20720	-	5.0000	7470A	PPB	05 Jun 2024 12:01:44	S	Std
7.5 - 1	57.5	29611	-	7.5000	7470A	PPB	05 Jun 2024 12:04:02	S	Std
10.0 - 1	510	39148	-	10.0000	7470A	PPB	05 Jun 2024 12:06:20	S	Std
ICV75 - 1	ICV75	16336	4.0787	-	7470A	PPB	05 Jun 2024 12:11:57	U	SMPL
ICB75 - 1	ICB75	281	-0.0305	-	7470A	PPB	05 Jun 2024 12:14:12	U	SMPL
CCV84 - 1	CCV84	19731	4.9476	-	7470A	PPB	05 Jun 2024 12:21:42	U	SMPL
CCB84 - 1	CCB84	230	-0.0435	-	7470A	PPB	05 Jun 2024 12:23:57	U	SMPL
HighStd - 1	HighStd	37519	9.5003	-	7470A	PPB	05 Jun 2024 12:28:30	U	SMPL
ChkStd - 1	ChkStd	25433	6.4070	-	7470A	PPB	05 Jun 2024 12:30:45	U	SMPL
CRA - 1	CRA	993	0.1518	-	7470A	PPB	05 Jun 2024 12:38:56	U	SMPL
PB161315BL - 1	PBW	319	-0.0208	-	7470A	PPB	05 Jun 2024 12:41:46	U	SMPL
PB161315BS - 1	LCSW	15968	3.9845	-	7470A	PPB	05 Jun 2024 12:46:39	U	SMPL
P2668-04 - 1	WC-TP-1	-117	-0.1323	-	7470A	PPB	05 Jun 2024 12:48:54	U	SMPL
P2668-08 - 1	WC-TP-2	38	-0.0927	-	7470A	PPB	05 Jun 2024 12:51:13	U	SMPL
P2681-01 - 1	SLUDGE-DRUM	88	-0.0799	-	7470A	PPB	05 Jun 2024 12:53:31	U	SMPL
P2683-03 - 1	WC-4	89	-0.0796	-	7470A	PPB	05 Jun 2024 12:55:48	U	SMPL
P2683-06 - 1	WC-5	65	-0.0858	-	7470A	PPB	05 Jun 2024 12:58:03	U	SMPL
CCV85 - 1	CCV85	20373	5.1119	-	7470A	PPB	05 Jun 2024 13:03:41	U	SMPL
CCB85 - 1	CCB85	278	-0.0312	-	7470A	PPB	05 Jun 2024 13:05:56	U	SMPL
P2686-04 - 1	WCS-TPQ	15	-0.0986	-	7470A	PPB	05 Jun 2024 13:08:13	U	SMPL
P2686-08 - 1	WCS-TPP	118	-0.0722	-	7470A	PPB	05 Jun 2024 13:10:29	U	SMPL
P2687-03 - 1	VNJ-223	63	-0.0863	-	7470A	PPB	05 Jun 2024 13:12:44	U	SMPL
P2687-06 - 1	72-11977	92	-0.0789	-	7470A	PPB	05 Jun 2024 13:15:00	U	SMPL
P2687-09 - 1	72-11998	102	-0.0763	-	7470A	PPB	05 Jun 2024 13:17:16	U	SMPL
P2687-09DUP - 1	72-11998DUP	85	-0.0806	-	7470A	PPB	05 Jun 2024 13:19:33	U	SMPL
P2687-09MS - 1	72-11998MS	18573	4.6512	-	7470A	PPB	05 Jun 2024 13:21:49	U	SMPL
P2687-09MSD - 1	72-11998MSD	18165	4.5468	-	7470A	PPB	05 Jun 2024 13:24:06	U	SMPL
PB161316BL - 1	PBW	138	-0.0671	-	7470A	PPB	05 Jun 2024 13:36:47	U	SMPL
PB161316BS - 1	LCSW	16751	4.1849	-	7470A	PPB	05 Jun 2024 13:39:02	U	SMPL
CCV86 - 1	CCV86	19325	4.8437	-	7470A	PPB	05 Jun 2024 13:41:18	U	SMPL
CCB86 - 1	CCB86	416	0.0041	-	7470A	PPB	05 Jun 2024 13:43:35	U	SMPL
P2691-02 - 1	RINSATE-1	5320	1.2592	-	7470A	PPB	05 Jun 2024 13:50:52	U	SMPL
P2691-02DUP - 1	RINSATE-1DUP	5106	1.2044	-	7470A	PPB	05 Jun 2024 13:53:07	U	SMPL
P2691-02MS - 1	RINSATE-1MS	19366	4.8542	-	7470A	PPB	05 Jun 2024 13:55:25	U	SMPL
P2691-02MSD - 1	RINSATE-1MSD	18260	4.5711	-	7470A	PPB	05 Jun 2024 14:00:39	U	SMPL
P2691-03 - 1	RINSATE-2	1235	0.2137	-	7470A	PPB	05 Jun 2024 14:02:55	U	SMPL
PB161353TB - 1	PB161353TB	548	0.0379	-	7470A	PPB	05 Jun 2024 14:07:50	U	SMPL
P2687-09LX5 - 1		-10	-0.1050	-	7470A	PPB	05 Jun 2024 14:10:06	U	SMPL
P2687-09A - 1		15746	3.9277	-	7470A	PPB	05 Jun 2024 14:12:23	U	SMPL
P2691-02LX5 - 1		1017	0.1579	-	7470A	PPB	05 Jun 2024 14:14:40	U	SMPL
P2691-02A - 1		20535	5.1534	-	7470A	PPB	05 Jun 2024 14:16:57	U	SMPL
CCV87 - 1	CCV87	18720	4.6889	-	7470A	PPB	05 Jun 2024 14:19:13	U	SMPL
CCB87 - 1	CCB87	341	-0.0151	-	7470A	PPB	05 Jun 2024 14:21:31	U	SMPL