

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB131080

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/05/2024	14:00
2	CAL2	1	Water	NA	NA	20.3	6.99	06/05/2024	14:01
3	CAL3	1	Water	NA	NA	20.3	10.02	06/05/2024	14:05
4	ICV	1	Water	NA	NA	20.4	7.01	06/05/2024	14:10
5	CCV1	1	Water	NA	NA	20.3	2.01	06/05/2024	14:15
6	P2740-04	1	Solid	20.03	20	23.6	5.75	06/05/2024	14:25
7	P2740-04DUP	1	Solid	20.04	20	23.7	5.76	06/05/2024	14:26
8	P2740-08	1	Solid	20.02	20	22.9	6.90	06/05/2024	14:30
9	CCV2	1	Water	NA	NA	20.3	12.02	06/05/2024	14:33

WORKLIST(Hardcopy Internal Chain)

W 131080

WorkList Name : corrosivity p2740 WorkList ID : 180816 Department : Wet-Chemistry Date : 06-05-2024 13:36:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2740-04	WCS-TPA	Solid	Corrosivity	Cool 4 deg C	PSEG03	M12	06/05/2024	9045D
P2740-08	WCS-TPB	Solid	Corrosivity	Cool 4 deg C	PSEG03	M12	06/05/2024	9045D

Date/Time 06/05/24 13:40
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 06/05/24 16:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]