

**Extraction and Analytical Summary Report**

Analysis Method: 1664A  
 Test: TPH  
 Run Number: LB131082  
 Analysis Date: 06/06/2024  
 BalanceID: WC SC-6  
 OvenID: EXT OVEN-3

ANALYST: jignesh  
 REVIEWED BY: Iwona  
 Extraction Date: 06/06/2024  
 Extration IN Time: 08:15  
 Extration OUT Time: 08:40  
 Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB131082BL	LB131082BL	WATER	1.3	1000	100	2.5634	2.5634	3.01	2.5635	2.5635	0.0001	0.1
2	LB131082BS	LB131082BS	WATER	1.3	1000	100	2.9633	2.9633	3.02	2.9800	2.9800	0.0167	16.7
3	LB131082BSD	LB131082BSD	WATER	1.3	1000	100	3.0251	3.0251	3.03	3.0420	3.0420	0.0169	16.9
4	P2713-01	402	WATER	1.3	1000	100	3.0158	3.0158	3.02	3.0160	3.0160	0.0002	0.2
5	P2714-01	N28660	WATER	1.6	1000	100	3.0131	3.0131	3.04	3.0149	3.0149	0.0018	1.8
6	P2722-01	Grab	WATER	1.6	1000	100	3.0184	3.0184	3.05	3.0867	3.0867	0.0683	68.3

QC Batch# LB131082Test: TPHAnalysis Date: 06/06/2024**Chemicals Used:**

Chemical Name	Chemical Lot #
HEXANE	W3076
pH Paper 0-14	M4909
Sodium Sulfate	EP2489
1:1 HCL	WP106917
Silica Gel	W3079
Sand	NA

**Standards Used:**

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP106918
LCSWD	5.00 ML	WP106919
MS/MSD	NA	NA

**BALANCE CALIBRATION / OVEN Dessicator Data****Analytical Balance ID # : WC SC-6****Before Analysis**

0.0020 gram Balance: 0.0018 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 10:26  
1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 09:15  
Bal Check Time: 08:30 Out OVEN TEMP1: 70 °C Dessicator Time Out1 11:00  
Out Time1 10:25

**After Analysis**

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 12:01  
1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time2: 11:30  
Bal Check Time: 12:40 Out OVEN TEMP2: 71 °C Dessicator Time Out2 12:37  
Out Time2: 12:00

VP 131082

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : tph p2722

WorkList ID : 180831

Department : Wet-Chemistry

Date : 06-06-2024 07:50:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2713-01	402	Water	TPH	Conc H2SO4 to pH < 2	PSEG04	K21	06/04/2024	1664A
P2714-01	N28660	Water	TPH	Conc H2SO4 to pH < 2	PSEG03	M21	06/04/2024	1664A
P2722-01	Grab	Water	TPH	Conc H2SO4 to pH < 2	ARAM01	K11	06/05/2024	1664A

Date/Time 06/06/24 08:10  
 Raw Sample Received by: JH CWC  
 Raw Sample Relinquished by: ACSM

Date/Time 06/06/24 12:40  
 Raw Sample Received by: ACSM  
 Raw Sample Relinquished by: JH CWC