


LB131086

SFAM01.1

INSTRUMENT ID: (V1

Linear



A= 0.0000e+000

B= 2.3432e-004 slope

C= 4.5399e-003 y-intercept

Rho= 0.9995754

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% 1)
0.00	0.000	0.013	0.013	38	0.000	38					11
0.05	0.050					442	0				12
0.20	0.200	0.197	-0.003	822	0.0 %	822					13
2.50	2.500	2.419	-0.081	10304	0.0 %	10304					14
5.00	5.000	5.192	0.192	22139	0.0 %	22139					12
7.50	7.500	7.344	-0.156	31320	0.0 %	31320					0
10.0	10.000	10.035	0.035	42804	0.0 %	42804					

LB131086

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	38	-		0 SFAM01.1	PPB	6/6/2024 10:38	Std
	0.2 S01	822	-		0.2 SFAM01.1	PPB	6/6/2024 10:40	Std
	2.5 S02	10304	-		2.5 SFAM01.1	PPB	6/6/2024 10:43	Std
	5 S03	22139	-		5 SFAM01.1	PPB	6/6/2024 10:45	Std
	7.5 S04	31320	-		7.5 SFAM01.1	PPB	6/6/2024 10:47	Std
	10 S05	42804	-		10 SFAM01.1	PPB	6/6/2024 10:50	Std
ICV077	ICV077	16228	3.8072 -		SFAM01.1	PPB	6/6/2024 10:52	SMPL
ICB077	ICB077	-42	-0.0053 -		SFAM01.1	PPB	6/6/2024 10:55	SMPL
CCV092	CCV092	22004	5.1606 -		SFAM01.1	PPB	6/6/2024 10:57	SMPL
CCB092	CCB092	-111	-0.0215 -		SFAM01.1	PPB	6/6/2024 10:59	SMPL
P2565-01DLX10	MX001	21713	5.0924 -		SFAM01.1	PPB	6/6/2024 11:01	SMPL
P2624-01DLX2	MBHBQ8	22137	5.1918 -		SFAM01.1	PPB	6/6/2024 11:04	SMPL
P2624-02DLX2	MBHBQ8D	21903	5.1369 -		SFAM01.1	PPB	6/6/2024 11:06	SMPL
P2624-03DLX2	MBHBQ8S	28559	6.6966 -		SFAM01.1	PPB	6/6/2024 11:08	SMPL
P2624-07DLX5	MBHBR2	11412	2.6786 -		SFAM01.1	PPB	6/6/2024 11:10	SMPL
P2624-08DLX5	MBHBR3	16722	3.9229 -		SFAM01.1	PPB	6/6/2024 11:13	SMPL
CCV093	CCV093	21699	5.0891 -		SFAM01.1	PPB	6/6/2024 11:15	SMPL
CCB093	CCB093	-184	-0.0386 -		SFAM01.1	PPB	6/6/2024 11:17	SMPL