

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131090

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/06/2024	10:25
2	CAL2	1	Water	NA	NA	20.3	7.00	06/06/2024	10:26
3	CAL3	1	Water	NA	NA	20.4	10.03	06/06/2024	10:30
4	ICV	1	Water	NA	NA	20.4	6.98	06/06/2024	10:35
5	CCV1	1	Water	NA	NA	20.5	2.02	06/06/2024	10:40
6	P2598-01	1	Water	NA	NA	20.9	7.06	06/06/2024	10:45
7	P2598-01DUP	1	Water	NA	NA	20.8	7.07	06/06/2024	10:46
8	P2598-02	1	Water	NA	NA	20.1	7.14	06/06/2024	10:50
9	P2598-03	1	Water	NA	NA	20.7	7.43	06/06/2024	10:55
10	P2723-01	1	Water	NA	NA	23.0	4.33	06/06/2024	11:00
11	P2723-02	1	Water	NA	NA	23.1	4.60	06/06/2024	11:05
12	P2723-03	1	Water	NA	NA	23.1	4.61	06/06/2024	11:08
13	CCV2	1	Water	NA	NA	20.7	12.02	06/06/2024	11:10

WORKLIST(Hardcopy Internal Chain)

UB131090

WorkList Name : PH P2598 WorkList ID : 180849 Department : Wet-Chemistry Date : 06-06-2024 10:03:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2598-01	SEEPAGE-CHAMBER	Water	pH	Cool 4 deg C	JPCL01	K21	06/04/2024	9040C
P2598-02	WETLANDS	Water	pH	Cool 4 deg C	JPCL01	K21	06/04/2024	9040C
P2598-03	I-VAULT	Water	pH	Cool 4 deg C	JPCL01	K21	06/04/2024	9040C
P2723-01	CELL-A-POOL-WATER	Water	pH	Cool 4 deg C	STER13	K11	06/04/2024	9040C
P2723-02	CELL-B-POOL-WATER	Water	pH	Cool 4 deg C	STER13	K11	06/04/2024	9040C
P2723-03	CELL-C-POOL-WATER	Water	pH	Cool 4 deg C	STER13	K11	06/04/2024	9040C

Date/Time 06/06/24 10:20
Raw Sample Received by: JH CWC1
Raw Sample Relinquished by: JH Sam

Date/Time 06/06/24 13:00
Raw Sample Received by: JH Sam
Raw Sample Relinquished by: JH WWC1