


L8131092

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.9337e-004

C= 1.7527e-002

Rho= 0.9995717

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(%)
0.0	0.000	0.026	0.026	42	0.000	42					1
0.2	0.200	0.199	-0.001	938	0.0 %	938					1
2.5	2.500	2.458	-0.042	12621	0.0 %	12621					2
5.0	5.000	5.130	0.130	26440	0.0 %	26440					2
7.5	7.500	7.295	-0.205	37634	0.0 %	37634					3
10.0	10.000	10.092	0.092	52100	0.0 %	52100					1

LB131092 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Jun 2024 12:27:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	42	-	0.0000	7471B	PPB	06 Jun 2024 12:28:49	S	Std
0.2 - 1	50.2	938	-	0.2000	7471B	PPB	06 Jun 2024 12:31:06	S	Std
2.5 - 1	52.5	12621	-	2.5000	7471B	PPB	06 Jun 2024 12:33:24	S	Std
5.0 - 1	55	26440	-	5.0000	7471B	PPB	06 Jun 2024 12:35:40	S	Std
7.5 - 1	57.5	37634	-	7.5000	7471B	PPB	06 Jun 2024 12:37:59	S	Std
10.0 - 1	5N	52100	-	10.0000	7471B	PPB	06 Jun 2024 12:43:01	S	Std
ICV77 - 1	ICV77	18574	3.6092	-	7471B	PPB	06 Jun 2024 12:45:46	U	SMPL
ICB77 - 1	ICB77	-130	-0.0076	-	7471B	PPB	06 Jun 2024 12:48:03	U	SMPL
CCV94 - 1	CCV94	26166	5.0773	-	7471B	PPB	06 Jun 2024 12:50:21	U	SMPL
CCB94 - 1	CCB94	-145	-0.0105	-	7471B	PPB	06 Jun 2024 12:52:36	U	SMPL
CRA - 1	CRA	922	0.1958	-	7471B	PPB	06 Jun 2024 12:54:56	U	SMPL
HighStd - 1	HighStd	51828	10.0396	-	7471B	PPB	06 Jun 2024 12:57:11	U	SMPL
ChkStd - 1	ChkStd	35105	6.8058	-	7471B	PPB	06 Jun 2024 12:59:26	U	SMPL
PB161355BL - 1	PBS	-274	-0.0355	-	7471B	PPB	06 Jun 2024 13:01:51	U	SMPL
PB161355BS - 1	LCSS	20563	3.9938	-	7471B	PPB	06 Jun 2024 13:04:13	U	SMPL
P2712-01 - 1	EO-01-060424	20490	3.9797	-	7471B	PPB	06 Jun 2024 13:06:30	U	SMPL
P2715-01 - 1	WCS-TPC	2167	0.4366	-	7471B	PPB	06 Jun 2024 13:08:49	U	SMPL
P2715-05 - 1	WCS-TPM	622	0.1378	-	7471B	PPB	06 Jun 2024 13:11:12	U	SMPL
P2740-01 - 1	WCS-TPA	1214	0.2523	-	7471B	PPB	06 Jun 2024 13:13:31	U	SMPL
P2740-05 - 1	WCS-TPB	1332	0.2751	-	7471B	PPB	06 Jun 2024 13:15:46	U	SMPL
CCV95 - 1	CCV95	25522	4.9528	-	7471B	PPB	06 Jun 2024 13:18:01	U	SMPL
CCB95 - 1	CCB95	-49	0.0081	-	7471B	PPB	06 Jun 2024 13:20:17	U	SMPL
P2740-05DUP - 1	WCS-TPBDUP	1653	0.3372	-	7471B	PPB	06 Jun 2024 13:22:35	U	SMPL
P2740-05MS - 1	WCS-TPBMS	25815	5.0094	-	7471B	PPB	06 Jun 2024 13:24:50	U	SMPL
P2740-05MSD - 1	WCS-TPBMSD	24641	4.7824	-	7471B	PPB	06 Jun 2024 13:27:07	U	SMPL
P2740-05LX5 - 1		-48	0.0082	-	7471B	PPB	06 Jun 2024 13:29:25	U	SMPL
P2740-05A - 1		21826	4.2381	-	7471B	PPB	06 Jun 2024 13:31:44	U	SMPL
CCV96 - 1	CCV96	25566	4.9613	-	7471B	PPB	06 Jun 2024 13:34:00	U	SMPL
CCB96 - 1	CCB96	-143	-0.0101	-	7471B	PPB	06 Jun 2024 13:36:20	U	SMPL