

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/7/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 17:05
In Date: 06/06/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 06/07/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB131100

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2588-23	VHBLK002	1	1.00	1.00	2.00	2.00	100.0	vhblk
P2753-01	DUP-1	2	1.18	8.42	9.6	8.23	83.7	
P2753-02	DUP-2	3	1.15	8.40	9.55	7.88	80.1	
P2756-01	OR-02-06062024	4	1.16	8.61	9.77	8.7	87.6	
P2756-02	OR-02-06062024	5	1.19	8.72	9.91	8.85	87.8	
P2757-01	PE-1	6	1.19	8.65	9.84	7.68	75.0	
P2757-02	PE-4	7	1.15	8.82	9.97	9.11	90.2	
P2757-03	DUP-1	8	1.16	8.45	9.61	8.43	86.0	
P2763-01	MH-SLUDGE	10	1.15	8.65	9.8	4.96	44.0	sludge sample
P2764-01	K-2211-ESSEX-SWITCHING-STATION	11	1.00	1.00	2.00	2.00	100.0	oil sample
P2764-02	H-1308-ESSEX-SWITCHING-STATION	12	1.00	1.00	2.00	2.00	100.0	oil sample
P2765-01	M-2239-HINCHMAN-SUBSTATION	13	1.00	1.00	2.00	2.00	100.0	oil sample
P2766-01	M-2239-FAIRLAWN-SWITCH	14	1.00	1.00	2.00	2.00	100.0	oil sample
P2769-01	JG001-JG007	9	1.18	8.53	9.71	8.66	87.7	
P2773-01	N-2266	15	1.00	1.00	2.00	2.00	100.0	oil sample
P2774-01	H-2313	16	1.00	1.00	2.00	2.00	100.0	oil sample

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$