

Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131113

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/07/2024	10:00
2	CAL2	1	Water	NA	NA	20.3	7.00	06/07/2024	10:04
3	CAL3	1	Water	NA	NA	20.3	10.02	06/07/2024	10:05
4	ICV	1	Water	NA	NA	20.3	7.01	06/07/2024	10:10
5	CCV1	1	Water	NA	NA	20.3	2.02	06/07/2024	10:11
6	P2760-01	1	Water	NA	NA	20.1	5.56	06/07/2024	10:15
7	P2760-01DUP	1	Water	NA	NA	20.2	5.57	06/07/2024	10:16
8	P2760-03	1	Water	NA	NA	20.4	6.50	06/07/2024	10:20
9	P2760-05	1	Water	NA	NA	20.3	6.42	06/07/2024	10:25
10	P2760-07	1	Water	NA	NA	20.1	6.56	06/07/2024	10:30
11	CCV2	1	Water	NA	NA	20.4	12.02	06/07/2024	10:35