


LB131114

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.9734e-004

C= 3.7953e-003

Rho= 0.9999056

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/6/11
0.0	0.000	-0.003	-0.003	-22	0.000	-22	0				
0.2	0.200	0.205	0.005	678	0.0 %	678					
2.5	2.500	2.567	0.067	8621	0.0 %	8621					
5.0	5.000	4.912	-0.088	16508	0.0 %	16508					
7.5	7.500	7.470	-0.030	25111	0.0 %	25111					
10.0	10.000	10.048	0.048	33780	0.0 %	33780					

LB131114 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Jun 2024 10:56:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-22	-	0.0000	7470A	PPB	07 Jun 2024 11:04:52	S	Std
0.2 - 1	50.2	678	-	0.2000	7470A	PPB	07 Jun 2024 11:07:09	S	Std
2.5 - 1	52.5	8621	-	2.5000	7470A	PPB	07 Jun 2024 11:09:26	S	Std
5.0 - 1	55	16508	-	5.0000	7470A	PPB	07 Jun 2024 11:11:43	S	Std
7.5 - 1	57.5	25111	-	7.5000	7470A	PPB	07 Jun 2024 11:14:01	S	Std
10.0 - 1	516	33780	-	10.0000	7470A	PPB	07 Jun 2024 11:19:01	S	Std
ICV78 - 1	ICV78	12469	3.7113	-	7470A	PPB	07 Jun 2024 11:21:52	U	SMPL
ICB78 - 1	ICB78	-271	-0.0768	-	7470A	PPB	07 Jun 2024 11:24:06	U	SMPL
CCV01 - 1	CCV01	15593	4.6402	-	7470A	PPB	07 Jun 2024 11:28:57	U	SMPL
CCB01 - 1	CCB01	-257	-0.0726	-	7470A	PPB	07 Jun 2024 11:31:12	U	SMPL
CRA - 1	CRA	614	0.1864	-	7470A	PPB	07 Jun 2024 11:36:06	U	SMPL
HighStd - 1	HighStd	33670	10.0151	-	7470A	PPB	07 Jun 2024 11:38:20	U	SMPL
ChkStd - 1	ChkStd	23293	6.9296	-	7470A	PPB	07 Jun 2024 11:40:36	U	SMPL
PB161356BL - 1	PBW	87	0.0297	-	7470A	PPB	07 Jun 2024 11:45:29	U	SMPL
PB161356BS - 1	LCSW	13464	4.0071	-	7470A	PPB	07 Jun 2024 11:47:44	U	SMPL
P2704-04 - 1	WCS-TPO	-280	-0.0795	-	7470A	PPB	07 Jun 2024 11:50:00	U	SMPL
P2704-08 - 1	WCS-TPU	-125	-0.0334	-	7470A	PPB	07 Jun 2024 11:52:19	U	SMPL
P2704-12 - 1	WCS-TPN	-85	-0.0215	-	7470A	PPB	07 Jun 2024 11:54:36	U	SMPL
P2715-04 - 1	WCS-TPC	-58	-0.0135	-	7470A	PPB	07 Jun 2024 11:56:53	U	SMPL
P2715-08 - 1	WCS-TPM	-34	-0.0063	-	7470A	PPB	07 Jun 2024 11:59:09	U	SMPL
CCV02 - 1	CCV02	16194	4.8189	-	7470A	PPB	07 Jun 2024 12:04:04	U	SMPL
CCB02 - 1	CCB02	-297	-0.0845	-	7470A	PPB	07 Jun 2024 12:06:19	U	SMPL
P2715-08DUP - 1	WCS-TPMDUP	-66	-0.0158	-	7470A	PPB	07 Jun 2024 12:08:36	U	SMPL
P2715-08MS - 1	WCS-TPMMS	12688	3.7764	-	7470A	PPB	07 Jun 2024 12:10:51	U	SMPL
P2715-08MSD - 1	WCS-TPMMSD	12989	3.8659	-	7470A	PPB	07 Jun 2024 12:13:06	U	SMPL
PB161376BL - 1	PBW	-244	-0.0688	-	7470A	PPB	07 Jun 2024 12:15:24	U	SMPL
PB161376BS - 1	LCSW	15867	4.7216	-	7470A	PPB	07 Jun 2024 12:17:42	U	SMPL
P2740-04 - 1	WCS-TPA	-235	-0.0661	-	7470A	PPB	07 Jun 2024 12:19:58	U	SMPL
P2740-08 - 1	WCS-TPB	-46	-0.0099	-	7470A	PPB	07 Jun 2024 12:22:16	U	SMPL
P2740-08DUP - 1	WCS-TPBDUP	-19	-0.0019	-	7470A	PPB	07 Jun 2024 12:24:33	U	SMPL
P2740-08MS - 1	WCS-TPBMS	13319	3.9640	-	7470A	PPB	07 Jun 2024 12:26:50	U	SMPL
P2740-08MSD - 1	WCS-TPBMSD	13630	4.0565	-	7470A	PPB	07 Jun 2024 12:29:06	U	SMPL
CCV03 - 1	CCV03	17098	5.0876	-	7470A	PPB	07 Jun 2024 12:35:20	U	SMPL
CCB03 - 1	CCB03	-228	-0.0640	-	7470A	PPB	07 Jun 2024 12:37:35	U	SMPL
PB161377BL - 1	PBW	-18	-0.0016	-	7470A	PPB	07 Jun 2024 12:39:52	U	SMPL
PB161377BS - 1	LCSW	12890	3.8365	-	7470A	PPB	07 Jun 2024 12:42:08	U	SMPL
P2714-01 - 1	N28660	-32	-0.0057	-	7470A	PPB	07 Jun 2024 12:44:24	U	SMPL
P2722-02 - 1	Comp	796	0.2405	-	7470A	PPB	07 Jun 2024 12:46:41	U	SMPL
P2749-01 - 1	SY-10S	-112	-0.0295	-	7470A	PPB	07 Jun 2024 12:48:57	U	SMPL
P2749-02 - 1	SY-10S	8	0.0062	-	7470A	PPB	07 Jun 2024 12:51:14	U	SMPL
P2749-03 - 1	SY-10D	95	0.0320	-	7470A	PPB	07 Jun 2024 12:53:31	U	SMPL
P2749-04 - 1	SY-10D	34	0.0139	-	7470A	PPB	07 Jun 2024 12:55:48	U	SMPL
P2749-05 - 1	SY-10I	1107	0.3329	-	7470A	PPB	07 Jun 2024 12:58:05	U	SMPL
P2749-06 - 1	SY-10I	815	0.2461	-	7470A	PPB	07 Jun 2024 13:00:21	U	SMPL
CCV04 - 1	CCV04	17930	5.3350	-	7470A	PPB	07 Jun 2024 13:16:29	U	SMPL
CCB04 - 1	CCB04	-261	-0.0738	-	7470A	PPB	07 Jun 2024 13:18:45	U	SMPL
P2749-07 - 1	SY-12D	5	0.0053	-	7470A	PPB	07 Jun 2024 13:21:02	U	SMPL
P2749-08 - 1	SY-12D	-4	0.0026	-	7470A	PPB	07 Jun 2024 13:23:18	U	SMPL
P2749-09 - 1	SY-12I	62	0.0222	-	7470A	PPB	07 Jun 2024 13:25:33	U	SMPL
P2749-09DUP - 1	SY-12IDUP	27	0.0118	-	7470A	PPB	07 Jun 2024 13:27:49	U	SMPL
P2749-09MS - 1	SY-12IMS	15313	4.5569	-	7470A	PPB	07 Jun 2024 13:30:05	U	SMPL
P2749-09MSD - 1	SY-12IMSD	13060	3.8870	-	7470A	PPB	07 Jun 2024 13:32:21	U	SMPL
P2749-10 - 1	SY-12I	-262	-0.0741	-	7470A	PPB	07 Jun 2024 13:34:40	U	SMPL
PB161296TB - 1	PB161296TB	-58	-0.0135	-	7470A	PPB	07 Jun 2024 13:36:59	U	SMPL
PB161338TB - 1	PB161338TB	-9	0.0011	-	7470A	PPB	07 Jun 2024 13:39:16	U	SMPL
CCV05 - 1	CCV05	16674	4.9616	-	7470A	PPB	07 Jun 2024 13:44:43	U	SMPL
CCB05 - 1	CCB05	-256	-0.0723	-	7470A	PPB	07 Jun 2024 13:46:58	U	SMPL
P2715-08LX5 - 1		8	0.0062	-	7470A	PPB	07 Jun 2024 13:49:16	U	SMPL
P2715-08A - 1		13394	3.9863	-	7470A	PPB	07 Jun 2024 13:51:32	U	SMPL
P2740-08LX5 - 1		-134	-0.0360	-	7470A	PPB	07 Jun 2024 13:53:47	U	SMPL
P2740-08A - 1		13760	4.0951	-	7470A	PPB	07 Jun 2024 13:56:05	U	SMPL
P2749-09LX5 - 1		-127	-0.0340	-	7470A	PPB	07 Jun 2024 13:58:21	U	SMPL
P2749-09A - 1		14228	4.2343	-	7470A	PPB	07 Jun 2024 14:00:38	U	SMPL
CCV06 - 1	CCV06	17340	5.1596	-	7470A	PPB	07 Jun 2024 14:05:30	U	SMPL
CCB06 - 1	CCB06	-239	-0.0673	-	7470A	PPB	07 Jun 2024 14:07:47	U	SMPL