

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB131115

ANALYST: Rubina
 SUPERVISOR REVIEW BY: Iwona
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W3105
IODINE SOLUTION .025N 1L	W3050
Starch Solution, 4L	W2977

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB161336BL		1	5.00	50	2.00	0.00	1.94	1.94	0.06	0.00	0.00	06/07/2024	10:45
2	P2704-04		1	5.05	50	2.00	0.00	1.92	1.92	0.08	0.02	1.58	06/07/2024	10:48
3	P2704-04DUP		1	5.04	50	2.00	0.00	1.92	1.92	0.08	0.02	1.59	06/07/2024	10:50
4	P2704-08		1	5.02	50	2.00	0.00	1.84	1.84	0.16	0.10	7.97	06/07/2024	10:53
5	P2704-12		1	5.05	50	2.00	0.00	1.88	1.88	0.12	0.06	4.75	06/07/2024	10:55
6	P2715-04		1	5.05	50	2.00	0.00	1.86	1.86	0.14	0.08	6.34	06/07/2024	10:58
7	P2715-08		1	5.03	50	2.00	0.00	1.92	1.92	0.08	0.02	1.59	06/07/2024	11:00
8	P2740-04		1	5.01	50	2.00	0.00	1.88	1.88	0.12	0.06	4.79	06/07/2024	11:03
9	P2740-08		1	5.03	50	2.00	0.00	1.86	1.86	0.14	0.08	6.36	06/07/2024	11:05
10	P2763-02		1	5.06	50	2.00	0.00	1.86	1.86	0.14	0.08	6.32	06/07/2024	11:08

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume