

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/10/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:50
In Date: 06/07/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 06/08/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131117

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2731-01	MYD3L6	1	1.12	8.73	9.85	9.7	98.3	
P2731-02	MYD3L7	2	1.16	8.50	9.66	9.45	97.5	
P2731-03	MYD3L8	3	1.12	8.77	9.89	9.75	98.4	
P2731-04	MYD3L9	4	1.16	8.48	9.64	9.42	97.4	
P2731-05	MYD3M0	5	1.15	8.34	9.49	9.26	97.2	
P2731-06	MYD3M1	6	1.16	8.61	9.77	9.6	98.0	
P2731-07	MYD3M3	7	1.15	8.40	9.55	9.41	98.3	
P2731-08	MYD3M5	8	1.19	8.42	9.61	9.47	98.3	
P2731-09	MYD3M6	9	1.19	8.67	9.86	9.74	98.6	
P2731-10	MYD3M7	10	1.15	8.79	9.94	9.8	98.4	
P2731-11	MYD3M8	11	1.11	8.81	9.92	9.7	97.5	
P2731-12	MYD3M9	12	1.19	8.38	9.57	9.45	98.6	
P2731-13	MYD3N0	13	1.12	8.82	9.94	9.65	96.7	
P2731-14	MYD3N1	14	1.13	8.38	9.51	9.35	98.1	
P2731-15	MYD3N2	15	1.15	8.75	9.9	9.63	96.9	
P2731-16	MYD3N3	16	1.16	8.66	9.82	9.57	97.1	
P2731-17	MYD3N3D	17	1.16	8.66	9.82	9.57	97.1	
P2731-18	MYD3N3S	18	1.16	8.66	9.82	9.57	97.1	
P2731-19	MYD3N4	19	1.15	8.45	9.6	9.47	98.5	
P2731-20	MYD3N5	20	1.12	8.46	9.58	9.4	97.9	
P2731-21	MYD3N6	21	1.17	8.55	9.72	9.55	98.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$