

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/10/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:20
In Date: 06/07/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 06/08/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131131

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2746-01	MYD3J4	1	1.17	8.55	9.72	9.55	98.0	
P2746-02	MYD3J4D	2	1.17	8.55	9.72	9.55	98.0	
P2746-03	MYD3J4S	3	1.17	8.55	9.72	9.55	98.0	
P2746-04	MYD403	4	1.16	8.58	9.74	9.5	97.2	
P2746-05	MYD3X4	5	1.16	8.48	9.64	9.51	98.5	
P2746-06	MYD3X5	6	1.15	8.84	9.99	9.67	96.4	
P2746-07	MYD3X6	7	1.13	8.54	9.67	9.57	98.8	
P2746-08	MYD3X7	8	1.12	8.45	9.57	9.37	97.6	
P2746-09	MYD3X8	9	1.14	8.80	9.94	9.63	96.5	
P2746-10	MYD3X9	10	1.15	8.35	9.5	9.18	96.2	
P2746-11	MYD3Y0	11	1.13	8.80	9.93	9.8	98.5	
P2746-12	MYD3Y1	12	1.13	8.53	9.66	9.41	97.1	
P2746-13	MYD3Y2	13	1.18	8.41	9.59	9.3	96.6	
P2746-14	MYD3Y3	14	1.16	8.62	9.78	9.62	98.1	
P2746-15	MYD3Y4	15	1.15	8.68	9.83	9.68	98.3	
P2746-16	MYD3Y5	16	1.16	8.65	9.81	9.7	98.7	
P2746-17	MYD3Y6	17	1.18	8.57	9.75	9.61	98.4	
P2746-18	MYD3Y7	18	1.19	8.53	9.72	9.61	98.7	
P2746-19	MYD3Y8	19	1.19	8.52	9.71	9.58	98.5	
P2746-20	MYD3Y9	20	1.16	8.50	9.66	9.41	97.1	
P2746-21	MYD3Z0	21	1.16	8.59	9.75	9.62	98.5	
P2746-22	MYD3Z1	22	1.11	8.74	9.85	9.72	98.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$