

LB131132


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40319



A= 0.0000e+000

B= 2.5150e-004 slope

C= -1.2482e-001 y-intercept

Rho= 0.9999208

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/d
0.00	0.000	0.019	0.019	571	0.000	571					=
0.05	0.050					442	0				23
0.20	0.200	0.245	0.045	1470	0.0 %	1470					-2
2.50	2.500	2.457	-0.043	10265	0.0 %	10265					-2
5.00	5.000	4.919	-0.081	20057	0.0 %	20057					-2
7.50	7.500	7.545	0.045	30495	0.0 %	30495					+
10.0	10.000	10.015	0.015	40319	0.0 %	40319					0

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	571	-		0 SFAM01.1	PPB	6/7/2024 18:30	Std
	0.2 S01	1470	-		0.2 SFAM01.1	PPB	6/7/2024 18:32	Std
	2.5 S02	10265	-		2.5 SFAM01.1	PPB	6/7/2024 18:35	Std
	5 S03	20057	-		5 SFAM01.1	PPB	6/7/2024 18:37	Std
	7.5 S04	30495	-		7.5 SFAM01.1	PPB	6/7/2024 18:41	Std
	10 S05	40319	-		10 SFAM01.1	PPB	6/7/2024 18:43	Std
ICV079	ICV079	16159	3.9392 -		SFAM01.1	PPB	6/7/2024 18:46	SMPL
ICB079	ICB079	537	0.0102 -		SFAM01.1	PPB	6/7/2024 18:48	SMPL
CCV007	CCV007	20049	4.9175 -		SFAM01.1	PPB	6/7/2024 18:51	SMPL
CCB007	CCB007	498	0.0004 -		SFAM01.1	PPB	6/7/2024 18:53	SMPL
PB161402BL	PBW402	498	0.0004 -		SFAM01.1	PPB	6/7/2024 18:58	SMPL
P2595-22	MBH5S1	18	-0.1203 -		SFAM01.1	PPB	6/7/2024 19:01	SMPL
P2625-16	MBH5S2	-4	-0.1258 -		SFAM01.1	PPB	6/7/2024 19:03	SMPL
P2650-10	BHBZ7	-3	-0.1256 -		SFAM01.1	PPB	6/7/2024 19:05	SMPL
P2650-18	BHBZ6	13	-0.1216 -		SFAM01.1	PPB	6/7/2024 19:07	SMPL
P2664-04	BHBZ5	5	-0.1236 -		SFAM01.1	PPB	6/7/2024 19:10	SMPL
P2697-01	MA4D61	60	-0.1097 -		SFAM01.1	PPB	6/7/2024 19:12	SMPL
P2697-19	MA4D66	25	-0.1185 -		SFAM01.1	PPB	6/7/2024 19:14	SMPL
P2709-01	GCKA0	1091	0.1496 -		SFAM01.1	PPB	6/7/2024 19:17	SMPL
P2709-02	GCKA0D	1109	0.1541 -		SFAM01.1	PPB	6/7/2024 19:19	SMPL
P2709-03	GCKA0S	5369	1.2255 -		SFAM01.1	PPB	6/7/2024 19:21	SMPL
P2709-04	GCKA1	1148	0.1639 -		SFAM01.1	PPB	6/7/2024 19:23	SMPL
P2709-05	GCKA2	1010	0.1292 -		SFAM01.1	PPB	6/7/2024 19:26	SMPL
P2709-06	GCKA3	974	0.1201 -		SFAM01.1	PPB	6/7/2024 19:28	SMPL
P2709-07	GCKA4	938	0.1111 -		SFAM01.1	PPB	6/7/2024 19:30	SMPL
P2709-08	GCKA5	1120	0.1569 -		SFAM01.1	PPB	6/7/2024 19:32	SMPL
P2709-09	GCKA6	1214	0.1805 -		SFAM01.1	PPB	6/7/2024 19:35	SMPL
P2709-10	GCKA7	1234	0.1855 -		SFAM01.1	PPB	6/7/2024 19:37	SMPL
P2709-11	GCKA8	1105	0.1531 -		SFAM01.1	PPB	6/7/2024 19:39	SMPL
P2709-12	GCKA9	230	-0.067 -		SFAM01.1	PPB	6/7/2024 19:41	SMPL
CCV008	CCV008	20711	5.084 -		SFAM01.1	PPB	6/7/2024 19:44	SMPL
CCB008	CCB008	569	0.0183 -		SFAM01.1	PPB	6/7/2024 19:46	SMPL
P2709-13	GCKB0	1049	0.139 -		SFAM01.1	PPB	6/7/2024 19:48	SMPL
PB161403BL	PBW403	477	-0.0049 -		SFAM01.1	PPB	6/7/2024 19:51	SMPL
P2708-01	GCK79	15	-0.121 -		SFAM01.1	PPB	6/7/2024 19:53	SMPL
P2708-02	GCK79D	4	-0.1238 -		SFAM01.1	PPB	6/7/2024 19:55	SMPL
P2708-03	GCK79S	3568	0.7725 -		SFAM01.1	PPB	6/7/2024 19:57	SMPL
P2708-04	GCK80	-69	-0.1422 -		SFAM01.1	PPB	6/7/2024 20:00	SMPL
P2708-05	GCK81	24	-0.1188 -		SFAM01.1	PPB	6/7/2024 20:02	SMPL
P2708-06	GCK82	0	-0.1248 -		SFAM01.1	PPB	6/7/2024 20:04	SMPL
P2708-07	GCK83	-1	-0.1251 -		SFAM01.1	PPB	6/7/2024 20:06	SMPL
P2708-08	GCK84	32	-0.1168 -		SFAM01.1	PPB	6/7/2024 20:09	SMPL
P2708-09	GCK85	2	-0.1243 -		SFAM01.1	PPB	6/7/2024 20:11	SMPL
P2708-10	GCK86	54	-0.1112 -		SFAM01.1	PPB	6/7/2024 20:13	SMPL
P2708-11	GCK87	13	-0.1216 -		SFAM01.1	PPB	6/7/2024 20:16	SMPL
P2708-12	GCK88	43	-0.114 -		SFAM01.1	PPB	6/7/2024 20:18	SMPL

LB131132

INSTRUMENT ID : CV1

P2708-13	GCK89	1368	0.2192 -	SFAM01.1	PPB	6/7/2024 20:20 SMPL
P2708-14	GCK90	1113	0.1551 -	SFAM01.1	PPB	6/7/2024 20:22 SMPL
CCV009	CCV009	20627	5.0629 -	SFAM01.1	PPB	6/7/2024 20:25 SMPL
CCB009	CCB009	163	-0.0838 -	SFAM01.1	PPB	6/7/2024 20:27 SMPL