


7471B

LB131151
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44689



A= 0.0000e+000

B= 2.2082e-004

C= -1.3568e-002

Rho= 0.9993777

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.068	0.068	369	0.000	369				
0.2	0.200	0.195	-0.005	944	0.0 %	944				
2.5	2.500	2.318	-0.182	10558	0.0 %	10558				
5.0	5.000	5.071	0.071	23025	0.0 %	23025				
7.5	7.500	7.694	0.194	34905	0.0 %	34905				
10.0	10.000	9.855	-0.145	44689	0.0 %	44689				

2/011
1-3
1-2
1-3
1-3

LB131151 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Jun 2024 10:51:37

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	369	-	0.000	7471B	PPB	10 Jun 2024 11:11:12	S	Std
0.2 - 1	50.2	944	-	0.200	7471B	PPB	10 Jun 2024 11:13:30	S	Std
2.5 - 1	52.5	10558	-	2.500	7471B	PPB	10 Jun 2024 11:15:47	S	Std
5.0 - 1	55	23025	-	5.000	7471B	PPB	10 Jun 2024 11:18:04	S	Std
7.5 - 1	57.5	34905	-	7.500	7471B	PPB	10 Jun 2024 11:20:24	S	Std
10.0 - 1	518	44689	-	10.000	7471B	PPB	10 Jun 2024 11:22:42	S	Std
ICV80 - 1	ICV80	18991	4.1800	-	7471B	PPB	10 Jun 2024 11:25:37	U	SMPL
ICB80 - 1	ICB80	354	0.0646	-	7471B	PPB	10 Jun 2024 11:27:53	U	SMPL
CCV10 - 1	CCV10	23902	5.2644	-	7471B	PPB	10 Jun 2024 11:30:11	U	SMPL
CCB10 - 1	CCB10	391	0.0728	-	7471B	PPB	10 Jun 2024 11:32:26	U	SMPL
CRA - 1	CRA	909	0.1872	-	7471B	PPB	10 Jun 2024 11:34:44	U	SMPL
HighStd - 1	HighStd	45538	10.0420	-	7471B	PPB	10 Jun 2024 11:36:59	U	SMPL
ChkStd - 1	ChkStd	33323	7.3447	-	7471B	PPB	10 Jun 2024 11:39:14	U	SMPL
PB161420BL - 1	PBS	-345	-0.0897	-	7471B	PPB	10 Jun 2024 11:41:40	U	SMPL
PB161420BS - 1	LCSS	19021	4.1866	-	7471B	PPB	10 Jun 2024 11:44:02	U	SMPL
P2756-01 - 1	OR-02-06062024	4373	0.9521	-	7471B	PPB	10 Jun 2024 11:46:18	U	SMPL
P2763-01 - 1	MH-SLUDGE	26195	5.7707	-	7471B	PPB	10 Jun 2024 11:48:38	U	SMPL
P2784-01 - 1	WCS-TPD	637	0.1271	-	7471B	PPB	10 Jun 2024 11:50:57	U	SMPL
P2786-01 - 1	HD-02-060724	3842	0.8348	-	7471B	PPB	10 Jun 2024 11:53:19	U	SMPL
P2787-01 - 1	SU-04-06072024	3160	0.6842	-	7471B	PPB	10 Jun 2024 11:55:34	U	SMPL
CCV11 - 1	CCV11	23246	5.1196	-	7471B	PPB	10 Jun 2024 11:57:51	U	SMPL
CCB11 - 1	CCB11	347	0.0631	-	7471B	PPB	10 Jun 2024 12:00:08	U	SMPL
P2787-01DUP - 1	SU-04-06072024DUP	3767	0.8183	-	7471B	PPB	10 Jun 2024 12:02:27	U	SMPL
P2787-01MS - 1	SU-04-06072024MS	22732	5.0061	-	7471B	PPB	10 Jun 2024 12:04:42	U	SMPL
P2787-01MSD - 1	SU-04-06072024MSD	22476	4.9495	-	7471B	PPB	10 Jun 2024 12:06:58	U	SMPL
P2794-01 - 1	TR-04-060724	2133	0.4574	-	7471B	PPB	10 Jun 2024 12:09:17	U	SMPL
P2795-01 - 1	WC-1-(0-5)-COMP	17669	3.8881	-	7471B	PPB	10 Jun 2024 12:11:36	U	SMPL
P2795-04 - 1	WC-2-(5-10)-COMP	3431	0.7441	-	7471B	PPB	10 Jun 2024 12:13:53	U	SMPL
P2795-07 - 1	WC-3-(10-15)-COMP	519	0.1010	-	7471B	PPB	10 Jun 2024 12:16:12	U	SMPL
P2787-01LX5 - 1		527	0.1028	-	7471B	PPB	10 Jun 2024 12:18:30	U	SMPL
P2787-01A - 1		22218	4.8926	-	7471B	PPB	10 Jun 2024 12:20:47	U	SMPL
CCV12 - 1	CCV12	23546	5.1858	-	7471B	PPB	10 Jun 2024 12:23:03	U	SMPL
CCB12 - 1	CCB12	388	0.0721	-	7471B	PPB	10 Jun 2024 12:25:22	U	SMPL