

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/11/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:50
In Date: 06/10/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:48
Out Date: 06/11/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131155

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2789-01	MYD424	1	1.16	8.50	9.66	9.38	96.7	
P2789-02	MYD425	2	1.15	8.36	9.51	9.36	98.2	
P2789-03	MYD425D	3	1.15	8.36	9.51	9.36	98.2	
P2789-04	MYD425S	4	1.15	8.36	9.51	9.36	98.2	
P2789-05	MYD426	5	1.19	8.56	9.75	9.54	97.5	
P2789-06	MYD427	6	1.19	8.42	9.61	9.5	98.7	
P2789-07	MYD428	7	1.13	8.80	9.93	9.77	98.2	
P2789-08	MYD429	8	1.16	8.67	9.83	9.68	98.3	
P2789-09	MYD431	9	1.18	8.49	9.67	9.55	98.6	
P2789-10	MYD432	10	1.15	8.40	9.55	9.12	94.9	
P2789-11	MYD433	11	1.16	8.50	9.66	9.51	98.2	
P2789-12	MYD434	12	1.18	8.53	9.71	9.55	98.1	
P2789-13	MYD435	13	1.18	8.51	9.69	9.56	98.5	
P2789-14	MYD436	14	1.15	8.37	9.52	9.4	98.6	
P2789-15	MYD439	15	1.14	8.68	9.82	9.66	98.2	
P2789-16	MYD440	16	1.18	8.40	9.58	9.38	97.6	
P2789-17	MYD441	17	1.18	8.35	9.53	9.36	98.0	
P2789-18	MYD443	18	1.15	8.66	9.81	9.6	97.6	
P2789-19	MYD444	19	1.19	8.60	9.79	9.66	98.5	
P2789-20	MYD4F7	20	1.12	8.52	9.64	9.57	99.2	
P2789-21	MYD4F8	21	1.15	8.40	9.55	9.41	98.3	
P2789-22	MYD4F9	22	1.16	8.68	9.84	9.7	98.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$