

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/11/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 14:15
In Date: 06/10/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:48
Out Date: 06/11/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131156

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2790-01	MYD445	1	1.18	8.40	9.58	9.41	98.0	
P2790-02	MYD446	2	1.16	8.50	9.66	9.47	97.8	
P2790-03	MYD447	3	1.18	8.59	9.77	9.58	97.8	
P2790-04	MYD447D	4	1.18	8.59	9.77	9.58	97.8	
P2790-05	MYD447S	5	1.18	8.59	9.77	9.58	97.8	
P2790-06	MYD448	6	1.18	8.76	9.94	9.59	96.0	
P2790-07	MYD449	7	1.16	8.41	9.57	9.42	98.2	
P2790-08	MYD450	8	1.19	8.57	9.76	9.43	96.1	
P2790-09	MYD451	9	1.14	8.39	9.53	9.38	98.2	
P2790-10	MYD452	10	1.12	8.54	9.66	9.5	98.1	
P2790-11	MYD453	11	1.15	8.40	9.55	9.44	98.7	
P2790-12	MYD454	12	1.19	8.43	9.62	9.38	97.2	
P2790-13	MYD4G0	13	1.14	8.39	9.53	9.44	98.9	
P2790-14	MYD4G1	14	1.18	8.70	9.88	9.6	96.8	
P2790-15	MYD4G2	15	1.13	8.45	9.58	9.47	98.7	
P2790-16	MYD4G3	16	1.18	8.37	9.55	9.41	98.3	
P2790-17	MYD4G4	17	1.18	8.39	9.57	9.3	96.8	
P2790-18	MYD4G5	18	1.19	8.59	9.78	9.55	97.3	
P2790-19	MYD4G7	19	1.15	8.65	9.8	9.66	98.4	
P2790-20	MYD4G8	20	1.18	8.48	9.66	9.54	98.6	
P2790-21	MYD4G9	21	1.17	8.49	9.66	9.55	98.7	
P2790-22	MYD4H0	22	1.19	8.56	9.75	9.66	98.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$