

LB13)160


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40327



A= 0.0000e+000

B= 2.5089e-004 slope

C= -8.4758e-002 y-intercept

Rho= 0.9999595

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/10
0.00	0.000	-0.027	-0.027	231	0.000	231					
0.05	0.050					442	0				
0.20	0.200	0.235	0.035	1276	0.0 %	1276					
2.50	2.500	2.529	0.029	10419	0.0 %	10419					
5.00	5.000	4.954	-0.046	20084	0.0 %	20084					
7.50	7.500	7.475	-0.025	30133	0.0 %	30133					
10.0	10.000	10.033	0.033	40327	0.0 %	40327					

0/10
18
1
1
0
0

LB131160

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	231	-		0 SFAM01.1	PPB	6/10/2024 13:43	Std
	0.2 S01	1276	-		0.2 SFAM01.1	PPB	6/10/2024 13:50	Std
	2.5 S02	10419	-		2.5 SFAM01.1	PPB	6/10/2024 13:53	Std
	5 S03	20084	-		5 SFAM01.1	PPB	6/10/2024 13:55	Std
	7.5 S04	30133	-		7.5 SFAM01.1	PPB	6/10/2024 13:57	Std
	10 S05	40327	-		10 SFAM01.1	PPB	6/10/2024 14:02	Std
ICV081	ICV081	16366	4.0213 -		SFAM01.1	PPB	6/10/2024 14:05	SMPL
ICB081	ICB081	124	-0.0536 -		SFAM01.1	PPB	6/10/2024 14:07	SMPL
CCV013	CCV013	20512	5.0615 -		SFAM01.1	PPB	6/10/2024 14:09	SMPL
CCB013	CCB013	99	-0.0599 -		SFAM01.1	PPB	6/10/2024 14:12	SMPL
PB161403BL	PBW403	150	-0.0471 -		SFAM01.1	PPB	6/10/2024 14:14	SMPL
P2708-01	GCK79	13	-0.0815 -		SFAM01.1	PPB	6/10/2024 14:16	SMPL
P2708-02	GCK79D	19	-0.08 -		SFAM01.1	PPB	6/10/2024 14:18	SMPL
P2708-03	GCK79S	3725	0.8498 -		SFAM01.1	PPB	6/10/2024 14:23	SMPL
P2708-04	GCK80	-33	-0.093 -		SFAM01.1	PPB	6/10/2024 14:26	SMPL
P2708-05	GCK81	54	-0.0712 -		SFAM01.1	PPB	6/10/2024 14:28	SMPL
P2708-06	GCK82	41	-0.0745 -		SFAM01.1	PPB	6/10/2024 14:30	SMPL
P2708-07	GCK83	28	-0.0777 -		SFAM01.1	PPB	6/10/2024 14:32	SMPL
P2708-08	GCK84	20	-0.0797 -		SFAM01.1	PPB	6/10/2024 14:35	SMPL
P2708-09	GCK85	24709	6.1144 -		SFAM01.1	PPB	6/10/2024 14:42	SMPL
P2708-10	GCK86	-307	-0.1618 -		SFAM01.1	PPB	6/10/2024 14:44	SMPL
P2708-11	GCK87	0	-0.0848 -		SFAM01.1	PPB	6/10/2024 14:46	SMPL
P2708-12	GCK88	1	-0.0845 -		SFAM01.1	PPB	6/10/2024 14:48	SMPL
P2708-13	GCK89	926	0.1476 -		SFAM01.1	PPB	6/10/2024 14:51	SMPL
P2708-14	GCK90	908	0.143 -		SFAM01.1	PPB	6/10/2024 14:53	SMPL
CCV014	CCV014	20669	5.1009 -		SFAM01.1	PPB	6/10/2024 14:55	SMPL
CCB014	CCB014	153	-0.0464 -		SFAM01.1	PPB	6/10/2024 14:58	SMPL