

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131165

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/10/2024	17:00
2	CAL2	1	Water	NA	NA	20.3	7.01	06/10/2024	17:01
3	CAL3	1	Water	NA	NA	20.2	10.02	06/10/2024	17:05
4	ICV	1	Water	NA	NA	20.3	7.00	06/10/2024	17:10
5	CCV1	1	Water	NA	NA	20.3	2.02	06/10/2024	17:15
6	P2802-01	1	Water	NA	NA	20.9	8.05	06/10/2024	17:20
7	P2802-01DUP	1	Water	NA	NA	20.8	8.06	06/10/2024	17:21
8	CCV2	1	Water	NA	NA	20.4	12.03	06/10/2024	17:22