

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/12/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:50
In Date: 06/11/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:28
Out Date: 06/12/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131171

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2818-01	MYD581	1	1.16	8.54	9.7	9.3	95.3	
P2818-02	MYD581D	2	1.16	8.54	9.7	9.3	95.3	
P2818-03	MYD581S	3	1.16	8.54	9.7	9.3	95.3	
P2818-04	MYD589	4	1.16	8.44	9.6	9.29	96.3	
P2818-05	MYD590	5	1.18	8.46	9.64	9.52	98.6	
P2818-06	MYD591	6	1.18	8.41	9.59	9.28	96.3	
P2818-07	MYD592	7	1.16	8.52	9.68	9.26	95.1	
P2818-08	MYD593	8	1.19	8.48	9.67	9.41	96.9	
P2818-09	MYD594	9	1.17	8.67	9.84	9.3	93.8	
P2818-10	MYD595	10	1.18	8.56	9.74	9.00	91.4	
P2818-11	MYD596	11	1.17	8.41	9.58	9.41	98.0	
P2818-12	MYD597	12	1.16	8.82	9.98	9.77	97.6	
P2818-13	MYD598	13	1.15	8.57	9.72	9.32	95.3	
P2818-14	MYD5A0	14	1.17	8.61	9.78	9.63	98.3	
P2818-15	MYD5A1	15	1.17	8.38	9.55	9.28	96.8	
P2818-16	MYD5A2	16	1.15	8.39	9.54	9.36	97.9	
P2818-17	MYD5A3	17	1.18	8.60	9.78	9.66	98.6	
P2818-18	MYD5A4	18	1.16	8.82	9.98	9.81	98.1	
P2818-19	MYD5A5	19	1.18	8.48	9.66	9.48	97.9	
P2818-20	MYD5A6	20	1.17	8.38	9.55	9.42	98.4	
P2818-21	MYD5A7	21	1.18	8.36	9.54	9.29	97.0	
P2818-22	MYD5A8	22	1.17	8.57	9.74	9.44	96.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$