

LB131172

7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43784

A= 0.0000e+000

B= 2.3109e-004

C= 5.2897e-002

Rho= 0.9988112

Accept=Accepted

Max. Conc: — 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.034	0.034	-81	0.000	-81	0				9.5
0.2	0.200	0.234	0.034	784	0.0 %	784					17
2.5	2.500	2.432	-0.068	10297	0.0 %	10297					3
5.0	5.000	5.182	0.182	22194	0.0 %	22194					4
7.5	7.500	7.147	-0.353	30696	0.0 %	30696					5
10.0	10.000	10.171	0.171	43784	0.0 %	43784					2

LB131172 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Jun 2024 10:02:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-81	-	0.0000	7470A	PPB	11 Jun 2024 10:10:38	S	Std
0.2 - 1	52.2	784	-	0.2000	7470A	PPB	11 Jun 2024 10:12:56	S	Std
2.5 - 1	55	10297	-	2.5000	7470A	PPB	11 Jun 2024 10:15:13	S	Std
5.0 - 1	57.5	22194	-	5.0000	7470A	PPB	11 Jun 2024 10:17:30	S	Std
7.5 - 1	57.5	30696	-	7.5000	7470A	PPB	11 Jun 2024 10:19:48	S	Std
10.0 - 1	57.5	43784	-	10.0000	7470A	PPB	11 Jun 2024 10:22:06	S	Std
ICV83 - 1	ICV83	17888	4.1867	-	7470A	PPB	11 Jun 2024 10:25:54	U	SMPL
ICB83 - 1	ICB83	-297	-0.0157	-	7470A	PPB	11 Jun 2024 10:28:08	U	SMPL
CCV18 - 1	CCV18	22059	5.1506	-	7470A	PPB	11 Jun 2024 10:30:26	U	SMPL
CCB18 - 1	CCB18	-299	-0.0162	-	7470A	PPB	11 Jun 2024 10:32:41	U	SMPL
CRA - 1	CRA	758	0.2281	-	7470A	PPB	11 Jun 2024 10:34:58	U	SMPL
HighStd - 1	HighStd	42183	9.8011	-	7470A	PPB	11 Jun 2024 10:37:14	U	SMPL
ChkStd - 1	ChkStd	28776	6.7028	-	7470A	PPB	11 Jun 2024 10:39:29	U	SMPL
PB161428BL - 1	PBW	-283	-0.0125	-	7470A	PPB	11 Jun 2024 10:41:47	U	SMPL
PB161428BS - 1	LCSW	16732	3.9195	-	7470A	PPB	11 Jun 2024 10:44:06	U	SMPL
P2763-02 - 1	MH-SLUDGE	-226	0.0007	-	7470A	PPB	11 Jun 2024 10:46:24	U	SMPL
P2763-02DUP - 1	MH-SLUDGEDUP	-146	0.0192	-	7470A	PPB	11 Jun 2024 10:48:42	U	SMPL
P2763-02MS - 1	MH-SLUDGEMS	14569	3.4197	-	7470A	PPB	11 Jun 2024 10:50:59	U	SMPL
P2763-02MSD - 1	MH-SLUDGEMSD	14103	3.3120	-	7470A	PPB	11 Jun 2024 10:53:16	U	SMPL
PB161429BL - 1	PBW	-514	-0.0659	-	7470A	PPB	11 Jun 2024 10:55:33	U	SMPL
CCV19 - 1	CCV19	20962	4.8971	-	7470A	PPB	11 Jun 2024 10:57:52	U	SMPL
CCB19 - 1	CCB19	-334	-0.0243	-	7470A	PPB	11 Jun 2024 11:00:08	U	SMPL
PB161429BS - 1	LCSW	15676	3.6755	-	7470A	PPB	11 Jun 2024 11:02:25	U	SMPL
P2784-04 - 1	WCS-TPD	-233	-0.0009	-	7470A	PPB	11 Jun 2024 11:04:41	U	SMPL
P2795-02 - 1	WC-1-(0-5)-COMP	181	0.0947	-	7470A	PPB	11 Jun 2024 11:06:59	U	SMPL
P2795-05 - 1	WC-2-(5-10)-COMP	-339	-0.0254	-	7470A	PPB	11 Jun 2024 11:09:15	U	SMPL
P2795-08 - 1	WC-3-(10-15)-COMP	-157	0.0166	-	7470A	PPB	11 Jun 2024 11:11:33	U	SMPL
P2795-08DUP - 1	WC-3-(10-15)-COMP DUP	-77	0.0351	-	7470A	PPB	11 Jun 2024 11:13:49	U	SMPL
P2795-08MS - 1	WC-3-(10-15)-COMP MS	13715	3.2223	-	7470A	PPB	11 Jun 2024 11:16:05	U	SMPL
P2795-08MSD - 1	WC-3-(10-15)-COMP MSD	13859	3.2556	-	7470A	PPB	11 Jun 2024 11:18:22	U	SMPL
CCV20 - 1	CCV20	20687	4.8335	-	7470A	PPB	11 Jun 2024 11:20:45	U	SMPL
CCB20 - 1	CCB20	-404	-0.0405	-	7470A	PPB	11 Jun 2024 11:23:02	U	SMPL
PB161387TB - 1		-106	0.0284	-	7470A	PPB	11 Jun 2024 11:25:20	U	SMPL
PB161401TB - 1		-63	0.0383	-	7470A	PPB	11 Jun 2024 11:27:35	U	SMPL
P2763-02LX5 - 1		-40	0.0437	-	7470A	PPB	11 Jun 2024 11:29:51	U	SMPL
P2763-02A - 1		15926	3.7333	-	7470A	PPB	11 Jun 2024 11:32:07	U	SMPL
P2795-08LX5 - 1		-180	0.0113	-	7470A	PPB	11 Jun 2024 11:34:23	U	SMPL
P2795-08A - 1		17334	4.0586	-	7470A	PPB	11 Jun 2024 11:36:40	U	SMPL
CCV21 - 1	CCV21	22320	5.2109	-	7470A	PPB	11 Jun 2024 11:38:56	U	SMPL
CCB21 - 1	CCB21	-357	-0.0296	-	7470A	PPB	11 Jun 2024 11:41:14	U	SMPL