

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/12/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:40
In Date: 06/11/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:37
Out Date: 06/12/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131175

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2810-01	MYD4G6	1	1.18	8.36	9.54	9.47	99.2	
P2810-02	MYD4G6D	2	1.18	8.36	9.54	9.47	99.2	
P2810-03	MYD4G6S	3	1.18	8.36	9.54	9.47	99.2	
P2810-04	MYD4S8	4	1.18	8.42	9.6	9.35	97.0	
P2810-05	MYD4S9	5	1.15	8.80	9.95	9.77	98.0	
P2810-06	MYD4T0	6	1.16	8.62	9.78	9.56	97.4	
P2810-07	MYD4T1	7	1.11	8.82	9.93	9.68	97.2	
P2810-08	MYD4T2	8	1.18	8.48	9.66	9.44	97.4	
P2810-09	MYD4T3	9	1.15	8.44	9.59	9.44	98.2	
P2810-10	MYD4T4	10	1.12	8.75	9.87	9.77	98.9	
P2810-11	MYD4T5	11	1.19	8.44	9.63	9.54	98.9	
P2810-12	MYD4T6	12	1.15	8.40	9.55	9.44	98.7	
P2810-13	MYD4T7	13	1.15	8.70	9.85	9.75	98.9	
P2810-14	MYD4T8	14	1.18	8.63	9.81	9.62	97.8	
P2810-15	MYD4T9	15	1.15	8.64	9.79	9.57	97.5	
P2810-16	MYD4W0	16	1.12	8.45	9.57	9.06	94.0	
P2810-17	MYD4W1	17	1.11	8.60	9.71	9.52	97.8	
P2810-18	MYD4W2	18	1.18	8.51	9.69	9.59	98.8	
P2810-19	MYD4W3	19	1.17	8.59	9.76	9.59	98.0	
P2810-20	MYD4W4	20	1.16	8.46	9.62	9.44	97.9	
P2810-21	MYD4W5	21	1.18	8.59	9.77	9.66	98.7	
P2810-22	MYD4W6	22	1.22	8.36	9.58	9.41	98.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$