

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/12/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 15:25
In Date: 06/11/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 06/12/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131181

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2814-01	MYD4H3	1	1.17	8.61	9.78	9.66	98.6	
P2814-02	MYD4H3D	2	1.17	8.61	9.78	9.66	98.6	
P2814-03	MYD4H3S	3	1.17	8.61	9.78	9.66	98.6	
P2814-04	MYD507	4	1.18	8.74	9.92	9.83	99.0	
P2814-05	MYD508	5	1.18	8.35	9.53	9.4	98.4	
P2814-06	MYD509	6	1.16	8.52	9.68	9.6	99.1	
P2814-07	MYD510	7	1.18	8.71	9.89	9.77	98.6	
P2814-08	MYD511	8	1.18	8.63	9.81	9.65	98.1	
P2814-09	MYD512	9	1.18	8.60	9.78	9.62	98.1	
P2814-10	MYD513	10	1.17	8.75	9.92	9.78	98.4	
P2814-11	MYD514	11	1.18	8.48	9.66	9.51	98.2	
P2814-12	MYD515	12	1.16	8.68	9.84	9.74	98.8	
P2814-13	MYD516	13	1.17	8.41	9.58	9.47	98.7	
P2814-14	MYD517	14	1.14	8.26	9.4	9.24	98.1	
P2814-15	MYD518	15	1.16	8.42	9.58	9.43	98.2	
P2814-16	MYD519	16	1.16	8.68	9.84	9.66	97.9	
P2814-17	MYD520	17	1.18	8.73	9.91	9.81	98.9	
P2814-18	MYD521	18	1.16	8.45	9.61	9.47	98.3	
P2814-19	MYD522	19	1.18	8.35	9.53	9.4	98.4	
P2814-20	MYD523	20	1.15	8.73	9.88	9.71	98.1	
P2814-21	MYD524	21	1.14	8.81	9.95	9.44	94.2	
P2814-22	MYD525	22	1.15	8.73	9.88	9.74	98.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$