

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/12/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:50
In Date: 06/11/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:12
Out Date: 06/12/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131184

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2815-01	MYD526	1	1.16	8.40	9.56	9.44	98.6	
P2815-02	MYD527	2	1.15	8.40	9.55	9.38	98.0	
P2815-03	MYD528	3	1.17	8.41	9.58	9.38	97.6	
P2815-04	MYD529	4	1.15	8.69	9.84	9.72	98.6	
P2815-05	MYD530	5	1.17	8.49	9.66	9.44	97.4	
P2815-06	MYD531	6	1.17	8.47	9.64	9.47	98.0	
P2815-07	MYD532	7	1.18	8.63	9.81	9.66	98.3	
P2815-08	MYD533	8	1.15	8.47	9.62	9.46	98.1	
P2815-09	MYD534	9	1.15	8.50	9.65	9.51	98.4	
P2815-10	MYD535	10	1.18	8.47	9.65	9.5	98.2	
P2815-11	MYD536	11	1.17	8.66	9.83	9.65	97.9	
P2815-12	MYD536D	12	1.17	8.66	9.83	9.65	97.9	
P2815-13	MYD536S	13	1.17	8.66	9.83	9.65	97.9	
P2815-14	MYD537	14	1.16	8.47	9.63	9.55	99.1	
P2815-15	MYD538	15	1.18	8.54	9.72	9.58	98.4	
P2815-16	MYD539	16	1.16	8.80	9.96	9.8	98.2	
P2815-17	MYD540	17	1.18	8.60	9.78	9.71	99.2	
P2815-18	MYD541	18	1.17	8.53	9.7	9.47	97.3	
P2815-19	MYD543	19	1.16	8.65	9.81	9.63	97.9	
P2815-20	MYD544	20	1.18	8.66	9.84	9.6	97.2	
P2815-21	MYD545	21	1.18	8.59	9.77	9.54	97.3	
P2815-22	MYD547	22	1.18	8.56	9.74	9.6	98.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$