

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131185

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	06/11/2024	15:10
2	P2843-01		1	11	100.00	>212.0	>212.0	06/11/2024	15:40
3	P2843-01DUP		1	11	100.00	>212.0	>212.0	06/11/2024	16:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6131185

WorkList Name : FP-6-11

WorkList ID : 180990

Department : Wet-Chemistry

Date : 06-11-2024 12:49:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2843-01	0416	Solid	Flash Point	Cool 4 deg C	PSEG03	M31	06/11/2024	1010B

Date/Time 06/11/2024 15.00
Raw Sample Received by: R.M. (WCS)
Raw Sample Relinquished by: R.M. (WCS)

Date/Time 06/11/2024 16.50
Raw Sample Received by: R.M. (WCS)
Raw Sample Relinquished by: R.M. (WCS)