

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/13/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:35
In Date: 06/12/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 06/13/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131187

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2816-01	MYD4J8	1	1.18	8.51	9.69	9.48	97.5	
P2816-02	MYD4J8D	2	1.18	8.51	9.69	9.48	97.5	
P2816-03	MYD4J8S	3	1.18	8.51	9.69	9.48	97.5	
P2816-04	MYD548	4	1.15	8.45	9.6	9.44	98.1	
P2816-05	MYD549	5	1.15	8.64	9.79	9.65	98.4	
P2816-06	MYD550	6	1.15	8.34	9.49	9.38	98.7	
P2816-07	MYD551	7	1.18	8.72	9.9	9.75	98.3	
P2816-08	MYD552	8	1.12	8.65	9.77	9.66	98.7	
P2816-09	MYD553	9	1.12	8.56	9.68	9.59	98.9	
P2816-10	MYD554	10	1.15	8.55	9.7	9.6	98.8	
P2816-11	MYD555	11	1.13	8.75	9.88	9.77	98.7	
P2816-12	MYD556	12	1.16	8.46	9.62	9.44	97.9	
P2816-13	MYD557	13	1.16	8.56	9.72	9.52	97.7	
P2816-14	MYD558	14	1.18	8.57	9.75	9.55	97.7	
P2816-15	MYD559	15	1.16	8.55	9.71	9.48	97.3	
P2816-16	MYD560	16	1.14	8.38	9.52	9.37	98.2	
P2816-17	MYD561	17	1.14	8.37	9.51	9.19	96.2	
P2816-18	MYD562	18	1.12	8.59	9.71	9.51	97.7	
P2816-19	MYD563	19	1.18	8.32	9.5	9.48	99.8	
P2816-20	MYD564	20	1.19	8.55	9.74	9.32	95.1	
P2816-21	MYD565	21	1.15	8.47	9.62	9.61	99.9	
P2816-22	MYD566	22	1.16	8.50	9.66	9.55	98.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$