

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/13/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:35
In Date: 06/12/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:55
Out Date: 06/13/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131192

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2821-01	MYD4L1	1	1.16	8.50	9.66	9.44	97.4	
P2821-02	MYD4L1D	2	1.16	8.50	9.66	9.44	97.4	
P2821-03	MYD4L1S	3	1.16	8.50	9.66	9.44	97.4	
P2821-04	MYD5E9	4	1.17	8.48	9.65	9.51	98.3	
P2821-05	MYD5F0	5	1.16	8.80	9.96	9.85	98.7	
P2821-06	MYD5F1	6	1.18	8.50	9.68	9.54	98.4	
P2821-07	MYD5F2	7	1.17	8.54	9.71	9.48	97.3	
P2821-08	MYD5F3	8	1.18	8.74	9.92	9.77	98.3	
P2821-09	MYD5F4	9	1.15	8.67	9.82	9.57	97.1	
P2821-10	MYD5F5	10	1.18	8.69	9.87	9.75	98.6	
P2821-11	MYD5F6	11	1.17	8.75	9.92	9.6	96.3	
P2821-12	MYD5F7	12	1.17	8.63	9.8	9.5	96.5	
P2821-13	MYD5F8	13	1.18	8.35	9.53	9.31	97.4	
P2821-14	MYD5F9	14	1.17	8.35	9.52	9.29	97.2	
P2821-15	MYD5G0	15	1.18	8.64	9.82	9.42	95.4	
P2821-16	MYD5G1	16	1.18	8.58	9.76	9.48	96.7	
P2821-17	MYD5G2	17	1.17	8.79	9.96	9.41	93.7	
P2821-18	MYD5G3	18	1.2	8.68	9.88	9.67	97.6	
P2821-19	MYD5G4	19	1.15	8.69	9.84	9.7	98.4	
P2821-20	MYD5G5	20	1.18	8.59	9.77	9.7	99.2	
P2821-21	MYD5G6	21	1.17	8.55	9.72	9.48	97.2	
P2821-22	MYD5G7	22	1.18	8.45	9.63	9.54	98.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$