

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/13/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 14:20
In Date: 06/12/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 06/13/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131195

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2822-01	MYD5H2	1	1.17	8.37	9.54	9.32	97.4	
P2822-02	MYD5H2D	2	1.17	8.37	9.54	9.32	97.4	
P2822-03	MYD5H2S	3	1.17	8.37	9.54	9.32	97.4	
P2822-04	MYD5H4	4	1.15	8.53	9.68	9.53	98.2	
P2822-05	MYD5H5	5	1.18	8.42	9.6	9.44	98.1	
P2822-06	MYD5H6	6	1.18	8.38	9.56	9.41	98.2	
P2822-07	MYD5H7	7	1.16	8.45	9.61	9.56	99.4	
P2822-08	MYD5H8	8	1.18	8.61	9.79	9.5	96.6	
P2822-09	MYD5H9	9	1.17	8.62	9.79	9.54	97.1	
P2822-10	MYD5J0	10	1.17	8.81	9.98	9.58	95.5	
P2822-11	MYD5J1	11	1.16	8.47	9.63	9.51	98.6	
P2822-12	MYD5J2	12	1.16	8.68	9.84	9.58	97.0	
P2822-13	MYD5J3	13	1.18	8.41	9.59	9.42	98.0	
P2822-14	MYD5J4	14	1.16	8.53	9.69	9.64	99.4	
P2822-15	MYD5J5	15	1.18	8.61	9.79	9.68	98.7	
P2822-16	MYD5J6	16	1.16	8.64	9.8	9.68	98.6	
P2822-17	MYD5J7	17	1.18	8.47	9.65	9.55	98.8	
P2822-18	MYD5J8	18	1.16	8.61	9.77	9.65	98.6	
P2822-19	MYD5K0	19	1.17	8.43	9.6	9.47	98.5	
P2822-20	MYD5K1	20	1.18	8.60	9.78	9.65	98.5	
P2822-21	MYD5K2	21	1.18	8.34	9.52	9.34	97.8	
P2822-22	MYD5K4	22	1.18	8.38	9.56	9.5	99.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$