

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/13/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 14:50
In Date: 06/12/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 06/13/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131196

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2825-01	MYD542	1	1.18	8.52	9.7	9.55	98.2	
P2825-02	MYD542D	2	1.18	8.52	9.7	9.55	98.2	
P2825-03	MYD542S	3	1.18	8.52	9.7	9.55	98.2	
P2825-04	MYD5P9	4	1.17	8.64	9.81	9.65	98.1	
P2825-05	MYD5Q0	5	1.15	8.40	9.55	9.3	97.0	
P2825-06	MYD5Q1	6	1.18	8.66	9.84	9.57	96.9	
P2825-07	MYD5Q2	7	1.18	8.57	9.75	9.55	97.7	
P2825-08	MYD5Q3	8	1.18	8.73	9.91	9.63	96.8	
P2825-09	MYD5Q4	9	1.18	8.37	9.55	9.27	96.7	
P2825-10	MYD5Q5	10	1.19	8.63	9.82	9.58	97.2	
P2825-11	MYD5Q6	11	1.18	8.59	9.77	9.54	97.3	
P2825-12	MYD5Q7	12	1.17	8.36	9.53	9.13	95.2	
P2825-13	MYD5Q8	13	1.18	8.63	9.81	9.44	95.7	
P2825-14	MYD5Q9	14	1.18	8.35	9.53	9.24	96.5	
P2825-15	MYD5R0	15	1.18	8.40	9.58	9.4	97.9	
P2825-16	MYD5R1	16	1.19	8.53	9.72	9.38	96.0	
P2825-17	MYD5R2	17	1.19	8.48	9.67	9.39	96.7	
P2825-18	MYD5G8	18	1.18	8.40	9.58	9.36	97.4	
P2825-19	MYD5G9	19	1.19	8.47	9.66	9.34	96.2	
P2825-20	MYD5H0	20	1.15	8.82	9.97	9.75	97.5	
P2825-21	MYD5H1	21	1.18	8.36	9.54	9.36	97.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$