

LB131197


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35742



A= 0.0000e+000

B= 2.7996e-004

C= -2.7303e-002

Rho= 0.9999534

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/10
0.0	0.000	-0.025	-0.025	10	0.000	10					76
0.05	0.050	0.047	-0.003	267	0.0 %	267	0				13
0.2	0.200	0.195	-0.005	795	0.0 %	795					3
2.5	2.500	2.563	0.063	9254	0.0 %	9254					3
5.0	5.000	4.951	-0.049	17781	0.0 %	17781					11
7.5	7.500	7.538	0.038	27024	0.0 %	27024					1
10.0	10.000	9.979	-0.021	35742	0.0 %	35742					0

LB131197 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 12 Jun 2024 14:02:40

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	10	-	0.0000	245.1	PPB	12 Jun 2024 14:12:32	S	Std
0.05 - 1	50.05	267	-	0.0500	245.1	PPB	12 Jun 2024 14:14:50	S	Std
0.2 - 1	50.2	795	-	0.2000	245.1	PPB	12 Jun 2024 14:17:07	S	Std
2.5 - 1	520.5	9254	-	2.5000	245.1	PPB	12 Jun 2024 14:19:24	S	Std
5.0 - 1	55	17781	-	5.0000	245.1	PPB	12 Jun 2024 14:21:40	S	Std
7.5 - 1	57.5	27024	-	7.5000	245.1	PPB	12 Jun 2024 14:23:58	S	Std
10.0 - 1	510	35742	-	10.0000	245.1	PPB	12 Jun 2024 14:31:41	S	Std
ICV85 - 1	ICV85	14952	4.1587	-245.1		PPB	12 Jun 2024 14:34:31	U	SMPL
ICB85 - 1	ICB85	-103	-0.0561	-245.1		PPB	12 Jun 2024 14:36:48	U	SMPL
CCV25 - 1	CCV25	18092	5.0378	-245.1		PPB	12 Jun 2024 14:39:05	U	SMPL
CCB25 - 1	CCB25	-100	-0.0553	-245.1		PPB	12 Jun 2024 14:41:20	U	SMPL
CRA - 1	CRA	271	0.0486	-245.1		PPB	12 Jun 2024 14:49:38	U	SMPL
HighStd - 1	HighStd	36145	10.0920	-245.1		PPB	12 Jun 2024 14:51:53	U	SMPL
ChkStd - 1	ChkStd	25208	7.0300	-245.1		PPB	12 Jun 2024 14:54:08	U	SMPL
PB161461BL - 1	PBW	194	0.0270	-245.1		PPB	12 Jun 2024 14:59:50	U	SMPL
PB161461BS - 1	LCSW	15093	4.1982	-245.1		PPB	12 Jun 2024 15:05:35	U	SMPL
P2403-10 - 1	LOD-MDL-WATER-04-QT2-2024	0	-0.0273	-245.1		PPB	12 Jun 2024 15:07:51	U	SMPL
P2403-10 - 1	LOD-MDL-WATER-04-QT2-2024	259	0.0452	-245.1		PPB	12 Jun 2024 15:10:11	U	SMPL
P2403-11 - 1	LOQ-WATER-05-QT2-2024	771	0.1885	-245.1		PPB	12 Jun 2024 15:12:27	U	SMPL
P2721-02 - 1	EFFLUENT-METALS	2117	0.5654	-245.1		PPB	12 Jun 2024 15:19:15	U	SMPL
P2721-02DUP - 1	EFFLUENT-METALSDUP	1943	0.5167	-245.1		PPB	12 Jun 2024 15:21:31	U	SMPL
P2721-02MS - 1	EFFLUENT-METALSMS	13376	3.7175	-245.1		PPB	12 Jun 2024 15:23:47	U	SMPL
CCV26 - 1	CCV26	17037	4.7424	-245.1		PPB	12 Jun 2024 15:27:56	U	SMPL
CCB26 - 1	CCB26	69	-0.0080	-245.1		PPB	12 Jun 2024 15:30:11	U	SMPL
P2721-02MSD - 1	EFFLUENT-METALSMSD	16051	4.4664	-245.1		PPB	12 Jun 2024 15:32:29	U	SMPL
P2721-02LX5 - 1		705	0.1701	-245.1		PPB	12 Jun 2024 16:07:19	U	SMPL
P2721-02A - 1		15106	4.2018	-245.1		PPB	12 Jun 2024 16:09:35	U	SMPL
CCV27 - 1	CCV27	17028	4.7399	-245.1		PPB	12 Jun 2024 16:11:50	U	SMPL
CCB27 - 1	CCB27	-59	-0.0438	-245.1		PPB	12 Jun 2024 16:14:09	U	SMPL