


LB131214
/INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.9140e-004

C= -9.2706e-002

Rho= 0.9996928

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Avg
0.0	0.000	-0.064	-0.064	150	0.000	150					150
0.2	0.200	0.194	-0.006	1497	0.0 %	1497					1497
2.5	2.500	2.492	-0.008	13504	0.0 %	13504					13504
5.0	5.000	5.101	0.101	27138	0.0 %	27138					27138
7.5	7.500	7.621	0.121	40304	0.0 %	40304					40304
10.0	10.000	9.855	-0.145	51975	0.0 %	51975					51975

LB131214 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Jun 2024 08:47:58

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	150	-	0.0000	7471B	PPB	13 Jun 2024 08:59:23	S	Std
0.2 - 1	50.2	1497	-	0.2000	7471B	PPB	13 Jun 2024 09:01:40	S	Std
2.5 - 1	52.5	13504	-	2.5000	7471B	PPB	13 Jun 2024 09:03:58	S	Std
5.0 - 1	55	27138	-	5.0000	7471B	PPB	13 Jun 2024 09:06:15	S	Std
7.5 - 1	57.5	40304	-	7.5000	7471B	PPB	13 Jun 2024 09:08:34	S	Std
10.0 - 1	59	51975	-	10.0000	7471B	PPB	13 Jun 2024 09:10:53	S	Std
ICV86 - 1	ICV86	21985	4.1152	-	7471B	PPB	13 Jun 2024 09:13:53	U	SMPL
ICB86 - 1	ICB86	27	-0.0875	-	7471B	PPB	13 Jun 2024 09:18:45	U	SMPL
CCV28 - 1	CCV28	27788	5.2259	-	7471B	PPB	13 Jun 2024 09:21:00	U	SMPL
CCB28 - 1	CCB28	316	-0.0322	-	7471B	PPB	13 Jun 2024 09:23:15	U	SMPL
CRA - 1	CRA	1507	0.1957	-	7471B	PPB	13 Jun 2024 09:25:33	U	SMPL
HighStd - 1	HighStd	52496	9.9550	-	7471B	PPB	13 Jun 2024 09:27:48	U	SMPL
ChkStd - 1	ChkStd	38846	7.3424	-	7471B	PPB	13 Jun 2024 09:30:04	U	SMPL
PB161464BL - 1	PB161464BL	186	-0.0571	-	7471B	PPB	13 Jun 2024 09:35:11	U	SMPL
PB161464BS - 1	PB161464BS	19815	3.6999	-	7471B	PPB	13 Jun 2024 09:37:27	U	SMPL
P2656-02 - 1	SB-12(1-4)	7070	1.2605	-	7471B	PPB	13 Jun 2024 09:39:44	U	SMPL
P2656-03 - 1	SB-13(1-4)	3999	0.6727	-	7471B	PPB	13 Jun 2024 09:42:04	U	SMPL
P2656-04 - 1	SB-14(0.2-2)	190908	36.4470	-	7471B	PPB	13 Jun 2024 09:44:24	U	SMPL
P2656-05 - 1	SB-15(1-3.5)	10755	1.9658	-	7471B	PPB	13 Jun 2024 09:46:42	U	SMPL
P2656-06 - 1	SB-16(0-2)	121866	23.2324	-	7471B	PPB	13 Jun 2024 09:49:19	U	SMPL
CCV29 - 1	CCV29	26523	4.9838	-	7471B	PPB	13 Jun 2024 09:51:37	U	SMPL
CCB29 - 1	CCB29	55	-0.0822	-	7471B	PPB	13 Jun 2024 09:54:06	U	SMPL
P2656-07 - 1	SB-17(0.5-2)	1012496	HIGH	-	7471B	PPB	13 Jun 2024 09:56:24	X	
P2656-08 - 1	SB-18(0.5-2)	212561	40.5913	-	7471B	PPB	13 Jun 2024 09:58:11	U	SMPL
P2656-09 - 1	SB-19(0-2)	1326352	HIGH	-	7471B	PPB	13 Jun 2024 10:01:08	X	
P2656-10 - 1	SB-20(0-2)	1142732	HIGH	-	7471B	PPB	13 Jun 2024 10:03:22	X	
P2656-17 - 1	DUP(X-X)	9827	1.7882	-	7471B	PPB	13 Jun 2024 10:06:25	U	SMPL
P2656-20 - 1	SB-11(1-4)	8665	1.5658	-	7471B	PPB	13 Jun 2024 10:09:44	U	SMPL
P2656-20DUP - 1	SB-11(1-4)DUP	8960	1.6222	-	7471B	PPB	13 Jun 2024 10:14:49	U	SMPL
P2656-21 - 1	P2656-20MS	28945	5.4474	-	7471B	PPB	13 Jun 2024 10:17:05	U	SMPL
P2656-22 - 1	P2656-20MSD	30331	5.7126	-	7471B	PPB	13 Jun 2024 10:19:25	U	SMPL
P2812-01 - 1	TR-05-061024	1823	0.2562	-	7471B	PPB	13 Jun 2024 10:21:44	U	SMPL
CCV30 - 1	CCV30	27400	5.1516	-	7471B	PPB	13 Jun 2024 10:24:05	U	SMPL
CCB30 - 1	CCB30	84	-0.0766	-	7471B	PPB	13 Jun 2024 10:26:23	U	SMPL
P2846-01 - 1	HR-03-061124	4872	0.8398	-	7471B	PPB	13 Jun 2024 10:28:42	U	SMPL
P2846-03 - 1	HR-04-061124	1715	0.2355	-	7471B	PPB	13 Jun 2024 10:30:57	U	SMPL
P2656-20LX5 - 1		1692	0.2311	-	7471B	PPB	13 Jun 2024 10:33:14	U	SMPL
P2656-20A - 1		29685	5.5890	-	7471B	PPB	13 Jun 2024 10:35:32	U	SMPL
P2656-04DLX10 - 1		17384	3.2346	-	7471B	PPB	13 Jun 2024 10:37:49	U	SMPL
P2656-06DLX5 - 1		23474	4.4002	-	7471B	PPB	13 Jun 2024 10:40:09	U	SMPL
P2656-07DLX10 - 1		182470	34.8319	-	7471B	PPB	13 Jun 2024 10:42:28	U	SMPL
P2656-08DLX10 - 1		17323	3.2229	-	7471B	PPB	13 Jun 2024 10:44:48	U	SMPL
P2656-09DLX10 - 1		394408	75.3968	-	7471B	PPB	13 Jun 2024 10:47:25	U	SMPL
P2656-10DLX10 - 1		252758	48.2850	-	7471B	PPB	13 Jun 2024 10:49:44	U	SMPL
CCV31 - 1	CCV31	27375	5.1469	-	7471B	PPB	13 Jun 2024 10:52:30	U	SMPL
CCB31 - 1	CCB31	74	-0.0785	-	7471B	PPB	13 Jun 2024 10:57:55	U	SMPL
P2656-07DL2X100 - 1		16009	2.9714	-	7471B	PPB	13 Jun 2024 11:00:11	U	SMPL
P2656-09DL2X100 - 1		43197	8.1752	-	7471B	PPB	13 Jun 2024 11:02:27	U	SMPL
P2656-10DL2X100 - 1		24772	4.6486	-	7471B	PPB	13 Jun 2024 11:04:45	U	SMPL
CCV32 - 1	CCV32	27412	5.1539	-	7471B	PPB	13 Jun 2024 11:07:07	U	SMPL
CCB32 - 1	CCB32	76	-0.0782	-	7471B	PPB	13 Jun 2024 11:09:26	U	SMPL