

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131216

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.00	06/13/2024	09:00
2	CAL2	1	Water	NA	NA	20.3	7.00	06/13/2024	09:01
3	CAL3	1	Water	NA	NA	20.3	10.02	06/13/2024	09:05
4	ICV	1	Water	NA	NA	20.2	7.01	06/13/2024	09:10
5	CCV1	1	Water	NA	NA	20.3	2.01	06/13/2024	09:15
6	P2854-02	1	Water	NA	NA	20.3	7.35	06/13/2024	09:25
7	P2854-02DUP	1	Water	NA	NA	20.4	7.36	06/13/2024	09:26
8	P2854-03	1	Water	NA	NA	20.7	6.87	06/13/2024	09:30
9	P2854-04	1	Water	NA	NA	20.7	7.38	06/13/2024	09:33
10	CCV2	1	Water	NA	NA	20.3	12.02	06/13/2024	09:37

WORKLIST(Hardcopy Internal Chain)

131216

WorkList Name : ph p2854

WorkList ID : 181029

Department : Wet-Chemistry

Date : 06-13-2024 07:50:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2854-02	60163	Water	pH	Cool 4 deg C	PSEG03	M31	06/12/2024	9040C
P2854-03	60164	Water	pH	Cool 4 deg C	PSEG03	M31	06/12/2024	9040C
P2854-04	60165	Water	pH	Cool 4 deg C	PSEG03	M31	06/12/2024	9040C

Date/Time 06.13.24 08:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 06.13.24 12:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]