

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/14/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:15
In Date: 06/13/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:35
Out Date: 06/14/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131218

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2824-01	MYD5K5	1	1.18	8.47	9.65	9.53	98.6	
P2824-02	MYD5K5D	2	1.18	8.47	9.65	9.53	98.6	
P2824-03	MYD5K5S	3	1.18	8.47	9.65	9.53	98.6	
P2824-04	MYD5N0	4	1.17	8.49	9.66	9.51	98.2	
P2824-05	MYD5N1	5	1.18	8.38	9.56	9.47	98.9	
P2824-06	MYD5N2	6	1.12	8.54	9.66	9.56	98.8	
P2824-07	MYD5N3	7	1.18	8.50	9.68	9.55	98.5	
P2824-08	MYD5N4	8	1.15	8.35	9.5	9.38	98.6	
P2824-09	MYD5N5	9	1.18	8.70	9.88	9.74	98.4	
P2824-10	MYD5N6	10	1.18	8.61	9.79	9.66	98.5	
P2824-11	MYD5N7	11	1.14	8.44	9.58	9.52	99.3	
P2824-12	MYD5N8	12	1.16	8.47	9.63	9.54	98.9	
P2824-13	MYD5N9	13	1.18	8.57	9.75	9.67	99.1	
P2824-14	MYD5P0	14	1.15	8.35	9.5	8.94	93.3	
P2824-15	MYD5P1	15	1.16	8.72	9.88	9.47	95.3	
P2824-16	MYD5P2	16	1.11	8.57	9.68	9.36	96.3	
P2824-17	MYD5P3	17	1.13	8.72	9.85	9.54	96.4	
P2824-18	MYD5P4	18	1.15	8.44	9.59	9.34	97.0	
P2824-19	MYD5P5	19	1.13	8.39	9.52	9.36	98.1	
P2824-20	MYD5P6	20	1.12	8.32	9.44	9.14	96.4	
P2824-21	MYD5P7	21	1.14	8.64	9.78	9.6	97.9	
P2824-22	MYD5P8	22	1.16	8.41	9.57	9.43	98.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$