

LB131227


SFAM01.1

INSTRUMENT ID (V)

Linear

 μ Abs.:

48578



A= 0.0000e+000

B= 2.0609e-004 *slope*C= 1.3637e-002 *intercept*

Rho= 0.9999582

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.00	0.000	0.043	0.043	144	0.000	144					
0.05	0.050					442	0				
0.20	0.200	0.210	0.010	951	0.0 %	951					
2.50	2.500	2.436	-0.064	11754	0.0 %	11754					
5.00	5.000	4.998	-0.002	24184	0.0 %	24184					
7.50	7.500	7.488	-0.012	36270	0.0 %	36270					
10.0	10.000	10.025	0.025	48578	0.0 %	48578					

9/11/12
5-30000

LB131227
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0		144	-		0 SFAM01.1	PPB	6/13/2024 14:59	Std
0.2		951	-		0.2 SFAM01.1	PPB	6/13/2024 15:01	Std
2.5		11754	-		2.5 SFAM01.1	PPB	6/13/2024 15:03	Std
5		24184	-		5 SFAM01.1	PPB	6/13/2024 15:06	Std
7.5		36270	-		7.5 SFAM01.1	PPB	6/13/2024 15:08	Std
10		48578	-		10 SFAM01.1	PPB	6/13/2024 15:12	Std
ICV087	ICV087	19134	3.9569 -		SFAM01.1	PPB	6/13/2024 15:15	SMPL
ICB087	ICB087	-45	0.0044 -		SFAM01.1	PPB	6/13/2024 15:17	SMPL
CCV033	CCV033	24327	5.0271 -		SFAM01.1	PPB	6/13/2024 15:19	SMPL
CCB033	CCB033	-60	0.0013 -		SFAM01.1	PPB	6/13/2024 15:22	SMPL
PB161481BL	PBS481	243	0.0637 -		SFAM01.1	PPB	6/13/2024 15:26	SMPL
P2650-01	MBHBT7	20031	4.1418 -		SFAM01.1	PPB	6/13/2024 15:29	SMPL
P2650-02	MBHBT8	21834	4.5134 -		SFAM01.1	PPB	6/13/2024 15:31	SMPL
P2650-03	MBHBT8D	18959	3.9209 -		SFAM01.1	PPB	6/13/2024 15:33	SMPL
P2650-05	MBHBT9	4629	0.9676 -		SFAM01.1	PPB	6/13/2024 15:38	SMPL
P2650-06	MBHBW0	3817	0.8003 -		SFAM01.1	PPB	6/13/2024 15:40	SMPL
P2650-07	MBHBW1	6538	1.361 -		SFAM01.1	PPB	6/13/2024 15:42	SMPL
P2650-08	MBHBW2	15984	3.3077 -		SFAM01.1	PPB	6/13/2024 15:45	SMPL
P2650-09	MBHBZ8	23407	4.8375 -		SFAM01.1	PPB	6/13/2024 15:47	SMPL
P2650-11	MBHC44	9236	1.9171 -		SFAM01.1	PPB	6/13/2024 15:49	SMPL
P2650-12	MBHC45	8940	1.8561 -		SFAM01.1	PPB	6/13/2024 15:51	SMPL
P2650-14	MBHC47	9558	1.9834 -		SFAM01.1	PPB	6/13/2024 15:54	SMPL
P2650-15	MBHC48	2957	0.623 -		SFAM01.1	PPB	6/13/2024 15:56	SMPL
P2650-16	MBHC50	7612	1.5824 -		SFAM01.1	PPB	6/13/2024 15:58	SMPL
P2650-17	MBHC51	8754	1.8177 -		SFAM01.1	PPB	6/13/2024 16:01	SMPL
PB161482BL	PBS482	-96	-0.0061 -		SFAM01.1	PPB	6/13/2024 16:03	SMPL
P2658-01	MBHC24	14924	3.0893 -		SFAM01.1	PPB	6/13/2024 16:05	SMPL
P2658-02	MBHC25	8718	1.8103 -		SFAM01.1	PPB	6/13/2024 16:07	SMPL
P2658-03	MBHC26	14186	2.9372 -		SFAM01.1	PPB	6/13/2024 16:10	SMPL
CCV034	CCV034	23335	4.8227 -		SFAM01.1	PPB	6/13/2024 16:12	SMPL
CCB034	CCB304	-93	-0.0055 -		SFAM01.1	PPB	6/13/2024 16:14	SMPL
P2658-04	MBHC26D	15276	3.1618 -		SFAM01.1	PPB	6/13/2024 16:17	SMPL
P2658-05	MBHC26S	27000	5.578 -		SFAM01.1	PPB	6/13/2024 16:19	SMPL
P2658-06	MBHC27	10763	2.2318 -		SFAM01.1	PPB	6/13/2024 16:21	SMPL
P2658-07	MBHC28	3108	0.6542 -		SFAM01.1	PPB	6/13/2024 16:23	SMPL
P2658-08	MBHC29	9622	1.9966 -		SFAM01.1	PPB	6/13/2024 16:26	SMPL
P2658-09	MBHC30	12534	2.5967 -		SFAM01.1	PPB	6/13/2024 16:28	SMPL
P2658-10	MBHC31	12576	2.6054 -		SFAM01.1	PPB	6/13/2024 16:30	SMPL
P2658-11	MBHC32	8831	1.8336 -		SFAM01.1	PPB	6/13/2024 16:32	SMPL
P2658-12	MBHC33	3000	0.6319 -		SFAM01.1	PPB	6/13/2024 16:35	SMPL
P2658-13	MBHC34	13000	2.6928 -		SFAM01.1	PPB	6/13/2024 16:37	SMPL
P2658-14	MBHC35	12813	2.6542 -		SFAM01.1	PPB	6/13/2024 16:39	SMPL
P2658-15	MBHC36	22185	4.5857 -		SFAM01.1	PPB	6/13/2024 16:42	SMPL
P2658-16	MBHC37	32413	6.6936 -		SFAM01.1	PPB	6/13/2024 16:44	SMPL
P2658-17	MBHC38	35399	7.3089 -		SFAM01.1	PPB	6/13/2024 16:46	SMPL
P2658-18	MBHC39	16256	3.3638 -		SFAM01.1	PPB	6/13/2024 16:48	SMPL

LB131227

INSTRUMENT ID : CV1

P2658-19	MBHC40	26549	5.4851 -	SFAM01.1	PPB	6/13/2024 16:51 SMPL
P2658-20	MBHC41	20971	4.3355 -	SFAM01.1	PPB	6/13/2024 16:53 SMPL
P2658-21	MBHC42	47908	9.8869 -	SFAM01.1	PPB	6/13/2024 16:55 SMPL
P2658-22	MBHC43	11841	2.4539 -	SFAM01.1	PPB	6/13/2024 16:58 SMPL
P2650-04	MBHBT8S	32471	6.7055 -	SFAM01.1	PPB	6/13/2024 17:00 SMPL
CCV035	CCV035	23648	4.8872 -	SFAM01.1	PPB	6/13/2024 17:02 SMPL
CCB035	CCB035	-114	-0.0099 -	SFAM01.1	PPB	6/13/2024 17:05 SMPL