

LB131240


7471B

INSTRUMENT ID: CVI

Linear

 μ Abs.:

55177



A= 0.0000e+000

B= 1.8035e-004

C= -3.3829e-002

Rho= 0.9996529

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.008	0.008	232	0.000	232					
0.2	0.200	0.193	-0.007	1257	0.0 %	1257					
2.5	2.500	2.376	-0.124	13363	0.0 %	13363					
5.0	5.000	5.185	0.185	28934	0.0 %	28934					
7.5	7.500	7.521	0.021	41888	0.0 %	41888					
10.0	10.000	9.918	-0.082	55177	0.0 %	55177					

0/60
 54
 55
 4
 0
 1

LB131240 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Jun 2024 11:18:24

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	232	-	0.0000	7471B	PPB	14 Jun 2024 11:36:56	S	Std
0.2 - 1	50.2	1257	-	0.2000	7471B	PPB	14 Jun 2024 11:39:13	S	Std
2.5 - 1	52.5	13363	-	2.5000	7471B	PPB	14 Jun 2024 11:41:32	S	Std
5.0 - 1	55	28934	-	5.0000	7471B	PPB	14 Jun 2024 11:43:50	S	Std
7.5 - 1	57.5	41888	-	7.5000	7471B	PPB	14 Jun 2024 11:46:10	S	Std
10.0 - 1	510	55177	-	10.0000	7471B	PPB	14 Jun 2024 11:48:29	S	Std
ICV89 - 1	ICV89	23527	4.2093	-	7471B	PPB	14 Jun 2024 11:51:22	U	SMPL
ICB89 - 1	ICB89	138	-0.0089	-	7471B	PPB	14 Jun 2024 11:53:39	U	SMPL
CCV39 - 1	CCV39	29642	5.3122	-	7471B	PPB	14 Jun 2024 11:55:57	U	SMPL
CCB39 - 1	CCB39	140	-0.0086	-	7471B	PPB	14 Jun 2024 11:58:11	U	SMPL
CRA - 1	CRA	1250	0.1916	-	7471B	PPB	14 Jun 2024 12:00:30	U	SMPL
HighStd - 1	HighStd	54740	9.8387	-	7471B	PPB	14 Jun 2024 12:02:45	U	SMPL
ChkStd - 1	ChkStd	38450	6.9008	-	7471B	PPB	14 Jun 2024 12:05:01	U	SMPL
PB161494BL - 1	PBS	-328	-0.0930	-	7471B	PPB	14 Jun 2024 12:07:25	U	SMPL
PB161494BS - 1	LCSS	23198	4.1500	-	7471B	PPB	14 Jun 2024 12:09:49	U	SMPL
P2855-05 - 1	72-11989	7705	1.3558	-	7471B	PPB	14 Jun 2024 12:12:06	U	SMPL
P2855-05DUP - 1	72-11989DUP	7054	1.2384	-	7471B	PPB	14 Jun 2024 12:14:25	U	SMPL
P2855-05MS - 1	72-11989MS	27468	4.9201	-	7471B	PPB	14 Jun 2024 12:16:45	U	SMPL
P2855-05MSD - 1	72-11989MSD	29445	5.2767	-	7471B	PPB	14 Jun 2024 12:19:04	U	SMPL
P2857-01 - 1	72-11985	123	-0.0116	-	7471B	PPB	14 Jun 2024 12:21:23	U	SMPL
CCV40 - 1	CCV40	28108	5.0355	-	7471B	PPB	14 Jun 2024 12:23:43	U	SMPL
CCB40 - 1	CCB40	-81	-0.0484	-	7471B	PPB	14 Jun 2024 12:25:59	U	SMPL
P2857-02 - 1	ETGI-355	869	0.1229	-	7471B	PPB	14 Jun 2024 12:28:17	U	SMPL
P2858-01 - 1	NB-07-06122024	1381	0.2152	-	7471B	PPB	14 Jun 2024 12:30:32	U	SMPL
P2859-01 - 1	OK-02-06122024	3536	0.6039	-	7471B	PPB	14 Jun 2024 12:32:48	U	SMPL
P2864-01 - 1	WC-H	1537	0.2434	-	7471B	PPB	14 Jun 2024 12:35:04	U	SMPL
P2864-05 - 1	WC-I	1723	0.2769	-	7471B	PPB	14 Jun 2024 12:37:22	U	SMPL
P2878-01 - 1	348	2845	0.4793	-	7471B	PPB	14 Jun 2024 12:39:38	U	SMPL
P2889-01 - 1	WCS-TPJ	790	0.1086	-	7471B	PPB	14 Jun 2024 12:41:55	U	SMPL
P2855-05LX5 - 1		1202	0.1830	-	7471B	PPB	14 Jun 2024 12:47:01	U	SMPL
P2855-05A - 1		27940	5.0052	-	7471B	PPB	14 Jun 2024 12:49:18	U	SMPL
CCV41 - 1	CCV41	28592	5.1228	-	7471B	PPB	14 Jun 2024 12:51:34	U	SMPL
CCB41 - 1	CCB41	-212	-0.0721	-	7471B	PPB	14 Jun 2024 12:53:53	U	SMPL