

LB 131249


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

41751



A= 0.0000e+000

B= 2.4150e-004

C= 3.5117e-002

Rho= 0.9995432

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	0.045	0.045	40	0.000	40	0				1
0.2	0.200	0.222	0.022	773	0.0 %	773					1
2.5	2.500	2.585	0.085	10559	0.0 %	10559					3
5.0	5.000	4.780	-0.220	19646	0.0 %	19646					4
7.5	7.500	7.451	-0.049	30707	0.0 %	30707					1
10.0	10.000	10.118	0.118	41751	0.0 %	41751					1

LB131243 IONSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Jun 2024 14:45:36

Sample ID	Extended ID	μ Abs	Conc	Std Conc Method	Units	Date	Type	Type
0.0 - 1	50	40	-	0.00007470A	PPB	14 Jun 2024 14:49:03	S	Std
0.2 - 1	51.2	773	-	0.20007470A	PPB	14 Jun 2024 14:51:20	S	Std
2.5 - 1	52.5	10559	-	2.50007470A	PPB	14 Jun 2024 14:53:37	S	Std
5.0 - 1	55	19648	-	5.00007470A	PPB	14 Jun 2024 15:02:39	S	Std
7.5 - 1	57.5	30707	-	7.50007470A	PPB	14 Jun 2024 15:07:43	S	Std
10.0 - 1	59	41751	-	10.00007470A	PPB	14 Jun 2024 15:12:38	S	Std
ICV90 - 1	ICV90	16168	3.9397	-7470A	PPB	14 Jun 2024 15:15:29	U	SMPL
ICB90 - 1	ICB90	-303	-0.0381	-7470A	PPB	14 Jun 2024 15:17:46	U	SMPL
CCV42 - 1	CCV42	19323	4.7016	-7470A	PPB	14 Jun 2024 15:20:04	U	SMPL
CCB42 - 1	CCB42	-406	-0.0629	-7470A	PPB	14 Jun 2024 15:22:19	U	SMPL
CRA - 1	CRA	666	0.1960	-7470A	PPB	14 Jun 2024 15:24:36	U	SMPL
HighStd - 1	HighStd	41072	9.9539	-7470A	PPB	14 Jun 2024 15:26:51	U	SMPL
ChkStd - 1	ChkStd	28418	6.8980	-7470A	PPB	14 Jun 2024 15:29:07	U	SMPL
PB161509BL - 1	PBW	-509	-0.0878	-7470A	PPB	14 Jun 2024 15:31:25	U	SMPL
PB161509BS - 1	LCSW	17065	4.1563	-7470A	PPB	14 Jun 2024 15:33:44	U	SMPL
P2856-19 - 1	RINSATE-BLANK	-71	0.0180	-7470A	PPB	14 Jun 2024 15:36:01	U	SMPL
P2893-02 - 1	TT-062-IDWGW-20240613	334	0.1158	-7470A	PPB	14 Jun 2024 15:38:20	U	SMPL
P2893-03 - 1	TT-063-IDWGW-20240613	176	0.0776	-7470A	PPB	14 Jun 2024 15:40:37	U	SMPL
P2893-04 - 1	TT-064-IDWGW-20240613	-108	0.0090	-7470A	PPB	14 Jun 2024 15:42:54	U	SMPL
P2893-04DUP - 1	TT-064-IDWGW-20240613DUP	-13	0.0320	-7470A	PPB	14 Jun 2024 15:45:09	U	SMPL
CCV43 - 1	CCV43	22158	5.3862	-7470A	PPB	14 Jun 2024 15:51:05	U	SMPL
CCB43 - 1	CCB43	-66	0.0192	-7470A	PPB	14 Jun 2024 15:55:55	U	SMPL
P2893-04MS - 1	TT-064-IDWGW-20240613MS	18131	4.4137	-7470A	PPB	14 Jun 2024 15:58:09	U	SMPL
P2893-04MSD - 1	TT-064-IDWGW-20240613MSD	16287	3.9684	-7470A	PPB	14 Jun 2024 16:00:25	U	SMPL
P2893-04LX5 - 1		-440	-0.0711	-7470A	PPB	14 Jun 2024 16:02:43	U	SMPL
P2893-04A - 1		17383	4.2331	-7470A	PPB	14 Jun 2024 16:05:02	U	SMPL
CCV44 - 1	CCV44	19537	4.7533	-7470A	PPB	14 Jun 2024 16:08:58	U	SMPL
CCB44 - 1	CCB44	-402	-0.0620	-7470A	PPB	14 Jun 2024 16:11:14	U	SMPL