

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB131249

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/15/2024	09:20
2	CAL2	1	Water	NA	NA	20.3	7.00	06/15/2024	09:21
3	CAL3	1	Water	NA	NA	20.3	10.03	06/15/2024	09:22
4	ICV	1	Water	NA	NA	20.4	7.01	06/15/2024	09:25
5	CCV1	1	Water	NA	NA	20.3	2.01	06/15/2024	09:30
6	P2878-03	1	Solid	20.02	20	20.4	8.17	06/15/2024	09:33
7	P2878-03DUP	1	Solid	20.03	20	20.5	8.18	06/15/2024	09:35
8	P2878-04	1	Solid	20.01	20	20.6	7.77	06/15/2024	09:40
9	P2889-04	1	Solid	20.02	20	21.3	8.25	06/15/2024	09:45
10	CCV2	1	Water	NA	NA	20.5	12.02	06/15/2024	09:49

WORKLIST(Hardcopy Internal Chain)

131249

WorkList Name : corrosivity p2889 WorkList ID : 181101 Department : Wet-Chemistry Date : 06-15-2024 09:07:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2878-03	348	Solid	Corrosivity	Cool 4 deg C	PSEG03	M41	06/13/2024	9045D
P2878-04	OILY-PADS	Solid	Corrosivity	Cool 4 deg C	PSEG03	M41	06/13/2024	9045D
P2889-04	WCS-TPJ	Solid	Corrosivity	Cool 4 deg C	PSEG03	M21	06/13/2024	9045D

Date/Time 06.15.24 09:10
Raw Sample Received by: SG(WOC)
Raw Sample Relinquished by: CSM

Date/Time 06.15.24 12:00
Raw Sample Received by: SG(WOC)
Raw Sample Relinquished by: SG(WOC)