

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/17/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:35
In Date: 06/15/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:40
Out Date: 06/16/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131252

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2845-01	MA4D90	1	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE
P2845-02	MA4C64	2	1.15	8.52	9.67	7.72	77.1	
P2845-03	MA4C64D	3	1.15	8.52	9.67	7.72	77.1	
P2845-04	MA4C64S	4	1.15	8.52	9.67	7.72	77.1	
P2845-05	MA4C46	5	1.18	8.42	9.6	8.74	89.8	
P2845-06	MA4C47	6	1.15	8.38	9.53	9.02	93.9	
P2845-07	MA4C48	7	1.18	8.34	9.52	9.14	95.4	
P2845-08	MA4C53	8	1.15	8.50	9.65	8.95	91.8	
P2845-09	MA4C57	9	1.18	8.42	9.6	8.78	90.3	
P2845-10	MA4CN4	10	1.18	8.33	9.51	6.45	63.3	
P2845-11	MA4CN5	11	1.19	8.53	9.72	7.36	72.3	
P2845-12	MA4CN7	12	1.17	8.55	9.72	8.92	90.6	
P2845-13	MA4C52	13	1.18	8.78	9.96	9.16	90.9	
P2845-14	MA4C54	14	1.12	8.77	9.89	6.71	63.7	
P2845-15	MA4C55	15	1.16	8.82	9.98	7.66	73.7	
P2845-16	MA4C58	16	1.15	8.75	9.9	9.09	90.7	
P2845-17	MA4C60	17	1.19	8.79	9.98	8.66	85.0	
P2845-18	MA4CJ1	18	1.16	8.69	9.85	5.08	45.1	
P2845-19	MA4CJ2	19	1.19	8.50	9.69	8.27	83.3	
P2845-20	MA4CN3	20	1.16	8.82	9.98	9.22	91.4	
P2845-21	MA4C49	21	1.18	8.65	9.83	8.6	85.8	
P2845-22	MA4C50	22	1.19	8.55	9.74	8.77	88.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$