

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/17/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:05
In Date: 06/15/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:40
Out Date: 06/16/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131254

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2900-01	EZWA6	1	1.18	8.57	9.75	8.81	89.0	
P2900-02	EZWA7	2	1.18	8.64	9.82	8.61	86.0	
P2900-03	EZWA8	3	1.18	8.69	9.87	8.76	87.2	
P2900-04	EZWA9	4	1.16	8.79	9.95	8.87	87.7	
P2900-05	EZWB0	5	1.16	8.83	9.99	8.72	85.6	
P2900-06	EZWB1	6	1.16	8.68	9.84	8.53	84.9	
P2900-07	EZWB2	7	1.16	8.48	9.64	8.2	83.0	
P2900-08	EZWB3	8	1.18	8.43	9.61	8.2	83.3	
P2900-09	EZWB4	9	1.18	8.37	9.55	7.9	80.3	
P2900-10	EZWB5	10	1.19	8.53	9.72	8.82	89.4	
P2900-11	EZWB6	11	1.19	8.66	9.85	8.89	88.9	
P2900-12	EZWB7	12	1.18	8.54	9.72	8.65	87.5	
P2900-13	EZWB8	13	1.18	8.74	9.92	8.86	87.9	
P2900-14	EZWB9	14	1.15	8.72	9.87	8.95	89.4	
P2900-15	EZWC0	15	1.16	8.70	9.86	9.00	90.1	
P2900-16	EZWC1	16	1.17	8.65	9.82	8.25	81.8	
P2900-17	EZWC2	17	1.18	8.45	9.63	8.19	83.0	
P2900-18	EZWC3	18	1.18	8.81	9.99	8.62	84.4	
P2900-19	EZWF1	19	1.16	8.75	9.91	8.56	84.6	
P2900-20	EZWF1D	20	1.16	8.75	9.91	8.56	84.6	
P2900-21	EZWF1S	21	1.16	8.75	9.91	8.56	84.6	
P2900-22	EZWF2	22	1.17	8.46	9.63	8.27	83.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$