

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/17/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:45
In Date: 06/15/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:40
Out Date: 06/16/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131257

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2835-01	MYD4A6	1	1.18	8.33	9.51	9.41	98.8	
P2835-02	MYD4A7	2	1.18	8.50	9.68	9.55	98.5	
P2835-03	MYD4A8	3	1.18	8.81	9.99	9.8	97.8	
P2835-04	MYD4A9	4	1.18	8.64	9.82	9.71	98.7	
P2835-05	MYD4A9D	5	1.18	8.64	9.82	9.71	98.7	
P2835-06	MYD4A9S	6	1.18	8.64	9.82	9.71	98.7	
P2835-07	MYD4B0	7	1.15	8.39	9.54	9.44	98.8	
P2835-08	MYD4B1	8	1.18	8.50	9.68	9.57	98.7	
P2835-09	MYD4B2	9	1.15	8.39	9.54	9.44	98.8	
P2835-10	MYD4B3	10	1.18	8.47	9.65	9.53	98.6	
P2835-11	MYD4B4	11	1.19	8.66	9.85	9.72	98.5	
P2835-12	MYD4B5	12	1.16	8.44	9.6	9.47	98.5	
P2835-13	MYD4B6	13	1.15	8.38	9.53	9.41	98.6	
P2835-14	MYD4B7	14	1.18	8.73	9.91	9.55	95.9	
P2835-15	MYD4B8	15	1.17	8.56	9.73	9.53	97.7	
P2835-16	MYD4B9	16	1.18	8.73	9.91	9.72	97.8	
P2835-17	MYD4C0	17	1.16	8.78	9.94	9.64	96.6	
P2835-18	MYD4C1	18	1.17	8.57	9.74	9.45	96.6	
P2835-19	MYD4C2	19	1.15	8.79	9.94	9.77	98.1	
P2835-20	MYD4C3	20	1.18	8.54	9.72	9.42	96.5	
P2835-21	MYD4C4	21	1.16	8.53	9.69	9.51	97.9	
P2835-22	MYD4C5	22	1.17	8.60	9.77	9.62	98.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$