

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/17/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:15
In Date: 06/15/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 06/16/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131258

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2836-01	MYD4C6	1	1.15	8.34	9.49	9.38	98.7	
P2836-02	MYD4C7	2	1.15	8.44	9.59	9.32	96.8	
P2836-03	MYD4C8	3	1.18	8.57	9.75	9.54	97.5	
P2836-04	MYD4C9	4	1.15	8.62	9.77	9.58	97.8	
P2836-05	MYD4D0	5	1.16	8.40	9.56	8.84	91.4	
P2836-06	MYD4D1	6	1.18	8.35	9.53	9.3	97.2	
P2836-07	MYD4D2	7	1.18	8.59	9.77	9.44	96.2	
P2836-08	MYD4D3	8	1.17	8.65	9.82	9.52	96.5	
P2836-09	MYD4D3D	9	1.17	8.65	9.82	9.52	96.5	
P2836-10	MYD4D3S	10	1.17	8.65	9.82	9.52	96.5	
P2836-11	MYD4D4	11	1.18	8.60	9.78	9.47	96.4	
P2836-12	MYD4D5	12	1.18	8.74	9.92	9.64	96.8	
P2836-13	MYD4D6	13	1.12	8.86	9.98	9.66	96.4	
P2836-14	MYD4D7	14	1.15	8.57	9.72	9.55	98.0	
P2836-15	MYD4D8	15	1.18	8.50	9.68	9.48	97.6	
P2836-16	MYD4D9	16	1.15	8.47	9.62	9.34	96.7	
P2836-17	MYD4E0	17	1.12	8.77	9.89	9.56	96.2	
P2836-18	MYD4E1	18	1.13	8.80	9.93	9.72	97.6	
P2836-19	MYD4E2	19	1.18	8.70	9.88	9.62	97.0	
P2836-20	MYD4E3	20	1.16	8.50	9.66	9.38	96.7	
P2836-21	MYD4E4	21	1.18	8.51	9.69	9.42	96.8	
P2836-22	MYD4E5	22	1.18	8.79	9.97	9.81	98.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$