

LB131259


7470A

INSTRUMENT ID: CV4

Linear

μ Abs.:

39515



A= 0.0000e+000

B= 2.5530e-004

C= 5.2024e-002

Rho= 0.9988742

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.057	0.057	20	0.000	20	0				0.01)
0.2	0.200	0.197	-0.003	567	0.0 %	567					-2
2.5	2.500	2.418	-0.082	9267	0.0 %	9267					-3
5.0	5.000	5.216	0.216	20227	0.0 %	20227					4
7.5	7.500	7.172	-0.328	27889	0.0 %	27889					-4
10.0	10.000	10.140	0.140	39515	0.0 %	39515					1

LB131259 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Jun 2024 09:10:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	20	-	0.0000	7470A	PPB	17 Jun 2024 09:54:03	S	Std
0.2 - 1	50-2	567	-	0.2000	7470A	PPB	17 Jun 2024 09:56:21	S	Std
2.5 - 1	52-5	9267	-	2.5000	7470A	PPB	17 Jun 2024 09:58:38	S	Std
5.0 - 1	55	20227	-	5.0000	7470A	PPB	17 Jun 2024 10:03:29	S	Std
7.5 - 1	58-5	27889	-	7.5000	7470A	PPB	17 Jun 2024 10:05:45	S	Std
10.0 - 1	510	39515	-	10.0000	7470A	PPB	17 Jun 2024 10:10:48	S	Std
ICV91 - 1	ICV91	15744	4.0715	-	7470A	PPB	17 Jun 2024 10:16:15	U	SMPL
ICB91 - 1	ICB91	-356	-0.0389	-	7470A	PPB	17 Jun 2024 10:18:30	U	SMPL
CCV45 - 1	CCV45	20663	5.3273	-	7470A	PPB	17 Jun 2024 10:20:47	U	SMPL
CCB45 - 1	CCB45	-359	-0.0396	-	7470A	PPB	17 Jun 2024 10:23:02	U	SMPL
CRA - 1	CRA	687	0.2274	-	7470A	PPB	17 Jun 2024 10:25:20	U	SMPL
HighStd - 1	HighStd	39038	10.0184	-	7470A	PPB	17 Jun 2024 10:27:38	U	SMPL
ChkStd - 1	ChkStd	26147	6.7273	-	7470A	PPB	17 Jun 2024 10:29:54	U	SMPL
PB161508BL - 1	PBW	-302	-0.0251	-	7470A	PPB	17 Jun 2024 10:32:12	U	SMPL
PB161508BS - 1	LCSW	16650	4.3028	-	7470A	PPB	17 Jun 2024 10:34:30	U	SMPL
P2853-01 - 1	1	41	0.0625	-	7470A	PPB	17 Jun 2024 10:36:47	U	SMPL
P2853-02 - 1	2	126	0.0842	-	7470A	PPB	17 Jun 2024 11:12:21	U	SMPL
P2853-03 - 1	3	14	0.0556	-	7470A	PPB	17 Jun 2024 11:14:44	U	SMPL
P2853-04 - 1	4	-118	0.0219	-	7470A	PPB	17 Jun 2024 11:17:01	U	SMPL
P2853-05 - 1	5	-294	-0.0230	-	7470A	PPB	17 Jun 2024 11:36:25	U	SMPL
CCV46 - 1	CCV46	20663	5.3273	-	7470A	PPB	17 Jun 2024 11:38:42	U	SMPL
CCB46 - 1	CCB46	-285	-0.0207	-	7470A	PPB	17 Jun 2024 11:40:58	U	SMPL
P2853-06 - 1	6	-102	0.0260	-	7470A	PPB	17 Jun 2024 11:43:16	U	SMPL
P2853-07 - 1	7	-43	0.0410	-	7470A	PPB	17 Jun 2024 11:45:31	U	SMPL
P2853-08 - 1	8	-14	0.0485	-	7470A	PPB	17 Jun 2024 11:47:47	U	SMPL
P2853-09 - 1	9	7	0.0538	-	7470A	PPB	17 Jun 2024 11:50:03	U	SMPL
P2853-10 - 1	10	-9	0.0497	-	7470A	PPB	17 Jun 2024 11:52:19	U	SMPL
P2853-11 - 1	11	1	0.0523	-	7470A	PPB	17 Jun 2024 11:54:35	U	SMPL
P2853-12 - 1	12	-4	0.0510	-	7470A	PPB	17 Jun 2024 11:56:52	U	SMPL
P2853-13 - 1	13	9	0.0543	-	7470A	PPB	17 Jun 2024 11:59:08	U	SMPL
P2864-04 - 1	WC-H	-2	0.0515	-	7470A	PPB	17 Jun 2024 12:01:26	U	SMPL
P2864-08 - 1	WC-I	19	0.0569	-	7470A	PPB	17 Jun 2024 12:03:41	U	SMPL
CCV47 - 1	CCV47	20486	5.2821	-	7470A	PPB	17 Jun 2024 12:05:57	U	SMPL
CCB47 - 1	CCB47	-259	-0.0141	-	7470A	PPB	17 Jun 2024 12:08:13	U	SMPL
P2878-03 - 1	348	-2	0.0515	-	7470A	PPB	17 Jun 2024 12:10:31	U	SMPL
P2889-04 - 1	WCS-TPJ	-8	0.0500	-	7470A	PPB	17 Jun 2024 12:12:47	U	SMPL
P2889-04DUP - 1	WCS-TPJDUP	-9	0.0497	-	7470A	PPB	17 Jun 2024 12:15:03	U	SMPL
P2889-04MS - 1	WCS-TPJMS	14970	3.8739	-	7470A	PPB	17 Jun 2024 12:17:18	U	SMPL
P2889-04MSD - 1	WCS-TPJMSD	13071	3.3890	-	7470A	PPB	17 Jun 2024 12:19:34	U	SMPL
PB161466TB - 1	PB161466TB	-201	0.0007	-	7470A	PPB	17 Jun 2024 12:21:52	U	SMPL
P2889-04LX5 - 1		16	0.0561	-	7470A	PPB	17 Jun 2024 12:24:11	U	SMPL
P2889-04A - 1		15440	3.9938	-	7470A	PPB	17 Jun 2024 12:26:28	U	SMPL
CCV48 - 1	CCV48	20767	5.3538	-	7470A	PPB	17 Jun 2024 12:28:45	U	SMPL
CCB48 - 1	CCB48	-96	0.0275	-	7470A	PPB	17 Jun 2024 12:31:01	U	SMPL