

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/18/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:05
In Date: 06/17/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:47
Out Date: 06/18/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131265

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2902-01	EZWD2	1	1.17	8.45	9.62	8.26	83.9	
P2902-02	EZWD3	2	1.18	8.56	9.74	8.49	85.4	
P2902-03	EZWD4	3	1.18	8.34	9.52	8.23	84.5	
P2902-04	EZWD5	4	1.19	8.53	9.72	8.16	81.7	
P2902-05	EZWD6	5	1.15	8.52	9.67	8.41	85.2	
P2902-06	EZWD7	6	1.16	8.67	9.83	8.53	85.0	
P2902-07	EZWD8	7	1.18	8.72	9.9	8.33	82.0	
P2902-08	EZWD9	8	1.18	8.45	9.63	8.08	81.7	
P2902-09	EZWE0	9	1.15	8.43	9.58	8.04	81.7	
P2902-10	EZWE1	10	1.12	8.74	9.86	8.21	81.1	
P2902-11	EZWE2	11	1.15	8.80	9.95	8.43	82.7	
P2902-12	EZWE3	12	1.19	8.52	9.71	8.06	80.6	
P2902-13	EZWE4	13	1.18	8.48	9.66	8.23	83.1	
P2902-14	EZWE5	14	1.16	8.42	9.58	8.07	82.1	
P2902-15	EZWE6	15	1.19	8.44	9.63	8.2	83.1	
P2902-16	EZWE7	16	1.17	8.74	9.91	8.52	84.1	
P2902-17	EZWE8	17	1.18	8.42	9.6	8.36	85.3	
P2902-18	EZWE9	18	1.17	8.60	9.77	8.54	85.7	
P2902-19	EZWFO	19	1.15	8.40	9.55	8.21	84.0	
P2902-20	EZWS9	20	1.18	8.48	9.66	8.33	84.3	
P2902-21	EZWS9D	21	1.18	8.48	9.66	8.33	84.3	
P2902-22	EZWS9S	22	1.18	8.48	9.66	8.33	84.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$