

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131268

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/17/2024	09:20
2	CAL2	1	Water	NA	NA	20.3	7.00	06/17/2024	09:21
3	CAL3	1	Water	NA	NA	20.3	10.03	06/17/2024	09:25
4	ICV	1	Water	NA	NA	20.3	7.02	06/17/2024	09:33
5	CCV1	1	Water	NA	NA	20.2	2.01	06/17/2024	09:35
6	P2879-01	1	Water	NA	NA	25.6	5.21	06/17/2024	09:40
7	P2893-02	1	Water	NA	NA	20.2	5.97	06/17/2024	09:47
8	P2893-03	1	Water	NA	NA	20.1	6.52	06/17/2024	09:50
9	P2893-04	1	Water	NA	NA	20.9	6.80	06/17/2024	09:55
10	P2893-04DUP	1	Water	NA	NA	20.8	6.81	06/17/2024	09:56
11	CCV2	1	Water	NA	NA	20.2	12.02	06/17/2024	10:00

WORKLIST(Hardcopy Internal Chain)

131268

WorkList Name : ph water p2893

WorkList ID : 181122

Department : Wet-Chemistry

Date : 06-17-2024 09:02:00

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2879-01	50363	Water	pH	Cool 4 deg C	PSEG03	M31	06/14/2024	9040C
P2893-02	TT-062-IDWGW-20240613	Water	pH	Cool 4 deg C	TETR06	M21	06/13/2024	9040C
P2893-03	TT-063-IDWGW-20240613	Water	pH	Cool 4 deg C	TETR06	M21	06/13/2024	9040C
P2893-04	TT-064-IDWGW-20240613	Water	pH	Cool 4 deg C	TETR06	M21	06/13/2024	9040C

Date/Time 06.17.24 09:15

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: