

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB131276

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/17/2024	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	06/17/2024	16:20
3	CAL3	1	Water	NA	NA	20.3	10.03	06/17/2024	16:21
4	ICV	1	Water	NA	NA	20.3	7.01	06/17/2024	16:25
5	CCV1	1	Water	NA	NA	20.2	2.01	06/17/2024	16:30
6	P2915-03	1	Solid	20.02	20	20.6	8.12	06/17/2024	16:33
7	P2915-03DUP	1	Solid	20.03	20	20.7	8.13	06/17/2024	16:34
8	P2915-04	1	Solid	20.04	20	20.9	9.02	06/17/2024	16:37
9	P2916-04	1	Solid	20.02	20	20.1	7.26	06/17/2024	16:42
10	P2917-03	1	Solid	20.03	20	20.3	5.49	06/17/2024	16:47
11	P2918-03	1	Solid	20.02	20	20.6	5.76	06/17/2024	16:52
12	P2926-01	1	Solid	20.03	20	20.9	8.46	06/17/2024	17:00
13	CCV2	1	Water	NA	NA	20.2	12.02	06/17/2024	17:02

✓B 131216

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity p2816 WorkList ID : 181135 Department : Wet-Chemistry Date : 06-17-2024 13:52:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2915-03	WC-NORTH-1	Solid	Corrosivity	Cool 4 deg C	PORT06	M31	06/14/2024	9045D
P2915-04	WC-NORTH-1	Solid	Corrosivity	Cool 4 deg C	PORT06	M31	06/14/2024	9045D
P2916-04	WCS-TPF	Solid	Corrosivity	Cool 4 deg C	PSEG03	M21	06/14/2024	9045D
P2917-03	ROAD	Solid	Corrosivity	Cool 4 deg C	PSEG03	M21	06/14/2024	9045D
P2918-03	POLE	Solid	Corrosivity	Cool 4 deg C	PSEG03	M21	06/14/2024	9045D
P2926-01	OILY-GRAVEL	Solid	Corrosivity	Cool 4 deg C	PSEG03	M21	06/17/2024	9045D

16:10

Date/Time 06-17-24
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 06-17-24
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]