


LB131282

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000
 B= 1.8394e-004 slope
 C= 2.3867e-002 y-intercept
 Rho= 0.9998937
 Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.046	0.046	118	0.000	118					=
0.05	0.050					442	0				17
0.20	0.200	0.233	0.033	1137	0.0 %	1137					-2
2.50	2.500	2.448	-0.052	13178	0.0 %	13178					-1
5.00	5.000	4.956	-0.044	26811	0.0 %	26811					-1
7.50	7.500	7.441	-0.059	40322	0.0 %	40322					-1
10.0	10.000	10.077	0.077	54656	0.0 %	54656					1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	118	-		0 SFAM01.1	PPB	6/18/2024 12:23	Std
	0.2 S01	1137	-		0.2 SFAM01.1	PPB	6/18/2024 12:26	Std
	2.5 S02	13178	-		2.5 SFAM01.1	PPB	6/18/2024 12:28	Std
	5 S03	26811	-		5 SFAM01.1	PPB	6/18/2024 12:30	Std
	7.5 S04	40322	-		7.5 SFAM01.1	PPB	6/18/2024 12:33	Std
	10 S05	54656	-		10 SFAM01.1	PPB	6/18/2024 12:35	Std
ICV093	ICV093	20651	3.8224 -		SFAM01.1	PPB	6/18/2024 12:38	SMPL
ICB093	ICB093	-75	0.0101 -		SFAM01.1	PPB	6/18/2024 12:40	SMPL
CCV051	CCV051	27325	5.0501 -		SFAM01.1	PPB	6/18/2024 12:43	SMPL
CCB051	CCB051	-144	-0.0026 -		SFAM01.1	PPB	6/18/2024 12:45	SMPL
PB161545BL	PBS545	20	0.0275 -		SFAM01.1	PPB	6/18/2024 12:47	SMPL
P2664-01	MBHC46	9709	1.8097 -		SFAM01.1	PPB	6/18/2024 12:50	SMPL
P2664-02	MBHC46D	10811	2.0125 -		SFAM01.1	PPB	6/18/2024 12:52	SMPL
P2664-05	MBH692	5792	1.0893 -		SFAM01.1	PPB	6/18/2024 12:57	SMPL
P2664-06	MBH693	5392	1.0157 -		SFAM01.1	PPB	6/18/2024 12:59	SMPL
P2664-07	MBH694	12947	2.4053 -		SFAM01.1	PPB	6/18/2024 13:01	SMPL
P2664-08	MBH695	7746	1.4487 -		SFAM01.1	PPB	6/18/2024 13:03	SMPL
P2664-09	MBH696	1958	0.384 -		SFAM01.1	PPB	6/18/2024 13:06	SMPL
P2664-10	MBH6B7	8292	1.5491 -		SFAM01.1	PPB	6/18/2024 13:08	SMPL
PB161546BL	PBS546	-135	-0.001 -		SFAM01.1	PPB	6/18/2024 13:10	SMPL
P2679-01	MBH672	12111	2.2516 -		SFAM01.1	PPB	6/18/2024 13:13	SMPL
P2679-02	MBH673	11576	2.1532 -		SFAM01.1	PPB	6/18/2024 13:15	SMPL
P2679-03	MBH674	18736	3.4702 -		SFAM01.1	PPB	6/18/2024 13:17	SMPL
P2679-04	MBH674D	17806	3.2991 -		SFAM01.1	PPB	6/18/2024 13:24	SMPL
P2679-06	MBH675	11434	2.127 -		SFAM01.1	PPB	6/18/2024 13:29	SMPL
P2679-07	MBH676	2482	0.4804 -		SFAM01.1	PPB	6/18/2024 13:31	SMPL
P2679-08	MBH677	11253	2.0938 -		SFAM01.1	PPB	6/18/2024 13:34	SMPL
P2679-09	MBH678	24194	4.4741 -		SFAM01.1	PPB	6/18/2024 13:36	SMPL
CCV052	CCV052	26269	4.8558 -		SFAM01.1	PPB	6/18/2024 13:38	SMPL
CCB052	CCB052	-221	-0.0168 -		SFAM01.1	PPB	6/18/2024 13:41	SMPL
P2679-10	MBH679	56087	10.3406 -		SFAM01.1	PPB	6/18/2024 13:43	SMPL
P2679-11	MBH680	10427	1.9418 -		SFAM01.1	PPB	6/18/2024 13:45	SMPL
P2679-12	MBH681	3793	0.7216 -		SFAM01.1	PPB	6/18/2024 13:48	SMPL
P2679-13	MBH682	16410	3.0423 -		SFAM01.1	PPB	6/18/2024 13:50	SMPL
P2679-14	MBH683	20018	3.706 -		SFAM01.1	PPB	6/18/2024 13:52	SMPL
P2679-15	MBH684	13999	2.5989 -		SFAM01.1	PPB	6/18/2024 13:54	SMPL
P2679-16	MBH685	9274	1.7297 -		SFAM01.1	PPB	6/18/2024 13:57	SMPL
P2679-17	MBH686	5987	1.1251 -		SFAM01.1	PPB	6/18/2024 13:59	SMPL
P2679-18	MBH687	7789	1.4566 -		SFAM01.1	PPB	6/18/2024 14:01	SMPL
P2679-19	MBH688	16062	2.9783 -		SFAM01.1	PPB	6/18/2024 14:04	SMPL
P2679-20	MBH689	11582	2.1543 -		SFAM01.1	PPB	6/18/2024 14:06	SMPL
P2679-21	MBH690	6465	1.213 -		SFAM01.1	PPB	6/18/2024 14:08	SMPL
P2679-22	MBH691	2397	0.4648 -		SFAM01.1	PPB	6/18/2024 14:11	SMPL
P2664-03	MBHC46S	23493	4.3452 -		SFAM01.1	PPB	6/18/2024 14:13	SMPL
P2679-05	MBH674S	32133	5.9344 -		SFAM01.1	PPB	6/18/2024 14:15	SMPL
P2679-10DLX2	MBH679	28012	5.1764 -		SFAM01.1	PPB	6/18/2024 14:18	SMPL

CCV053	CCV053	26540	4.9057 -	SFAM01.1	PPB	6/18/2024 14:20	SMPL
CCB053	CCB053	-156	-0.0048 -	SFAM01.1	PPB	6/18/2024 14:22	SMPL