

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB131286

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/18/2024	15:35
2	CAL2	1	Water	NA	NA	20.3	7.01	06/18/2024	15:36
3	CAL3	1	Water	NA	NA	20.3	10.02	06/18/2024	15:38
4	ICV	1	Water	NA	NA	20.2	6.98	06/18/2024	15:40
5	CCV1	1	Water	NA	NA	20.3	2.01	06/18/2024	15:44
6	P2930-04	1	Solid	20.02	20	21.6	6.22	06/18/2024	15:46
7	P2938-01	1	Solid	20.01	20	23.4	8.04	06/18/2024	15:50
8	P2938-02	1	Solid	20.03	20	24.3	7.14	06/18/2024	15:53
9	P2938-03	1	Solid	20.04	8.38	1600	23.3	06/18/2024	16:00
10	P2953-01	1	Solid	20.01	20.2	20.2	5.53	06/18/2024	16:05
11	P2953-02	1	Solid	20.02	20.2	20.3	5.51	06/18/2024	16:15
12	P2953-03	1	Solid	20.03	20.2	23.4	5.83	06/18/2024	16:20
13	P2955-04	1	Solid	20.02	20.2	23.6	9.34	06/18/2024	16:30
14	P2955-08	1	Solid	20.03	20.2	22.4	8.02	06/18/2024	16:35
15	P2955-08DUP	1	Solid	20.04	20.2	22.5	8.03	06/18/2024	16:36
16	CCV2	1	Water	NA	NA	20.2	12.02	06/18/2024	16:38

WORKLIST(Hardcopy Internal Chain)

VB 131286

WorkList Name : corrosivity p2930 SOLIDS WorkList ID : 181175 Department : Wet-Chemistry Date : 06-18-2024 15:00:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2930-04	WCS-TPE	Solid	Corrosivity	Cool 4 deg C	PSEG03	M21	06/17/2024	9045D
P2938-01	397	Solid	Corrosivity	Cool 4 deg C	PSEG03	M21	06/18/2024	9045D
P2938-02	398	Solid	Corrosivity	Cool 4 deg C	PSEG03	M21	06/18/2024	9045D
P2938-03	404	Solid	Corrosivity	Cool 4 deg C	PSEG03	M21	06/18/2024	9045D
P2953-01	SOM-24-00054	Solid	Corrosivity	Cool 4 deg C	PSEG03	M21	06/18/2024	9045D
P2953-02	COMP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	M21	06/18/2024	9045D
P2953-03	COMP-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	M21	06/18/2024	9045D
P2955-04	VK-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	M21	06/18/2024	9045D
P2955-08	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	M21	06/18/2024	9045D

Date/Time 06-18-24 15:30
Raw Sample Received by: Jof WWC
Raw Sample Relinquished by: Jof WWC

Date/Time 06-18-24 18:00
Raw Sample Received by: Jof WWC
Raw Sample Relinquished by: Jof WWC