

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131287

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/18/2024	15:35
2	CAL2	1	Water	NA	NA	20.3	7.01	06/18/2024	15:36
3	CAL3	1	Water	NA	NA	20.3	10.02	06/18/2024	15:38
4	ICV	1	Water	NA	NA	20.2	6.98	06/18/2024	15:40
5	CCV1	1	Water	NA	NA	20.2	2.02	06/18/2024	16:43
6	P2938-04	1	Water	NA	NA	25.6	6.95	06/18/2024	16:48
7	P2938-04DUP	1	Water	NA	NA	25.8	6.96	06/18/2024	16:49
8	P2938-05	1	Water	NA	NA	24.3	7.39	06/18/2024	16:55
9	P2938-06	1	Water	NA	NA	25.3	7.05	06/18/2024	17:00
10	P2938-07	1	Water	NA	NA	23.8	7.49	06/18/2024	17:05
11	P2942-01	1	Water	NA	NA	20.5	10.25	06/18/2024	17:15
12	CCV2	1	Water	NA	NA	20.3	12.02	06/18/2024	17:20

WORKLIST(Hardcopy Internal Chain)

131287

WorkList Name : PH P2938

WorkList ID : 181176

Department : Wet-Chemistry

Date : 06-18-2024 15:01:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2938-04	399-400	Water	pH	Cool 4 deg C	PSEG03	M21	06/18/2024	9040C
P2938-05	401-402	Water	pH	Cool 4 deg C	PSEG03	M21	06/18/2024	9040C
P2938-06	403	Water	pH	Cool 4 deg C	PSEG03	M21	06/18/2024	9040C
P2938-07	405	Water	pH	Cool 4 deg C	PSEG03	M21	06/18/2024	9040C
P2942-01	GRAY-WATER	Water	pH	Cool 4 deg C	PSEG03	M21	06/18/2024	9040C

Date/Time

06-18-24 15:10

Raw Sample Received by:

70 CWC

Raw Sample Relinquished by:

70 CWC

Date/Time

06-18-24

Raw Sample Received by:

70 CWC

Raw Sample Relinquished by: